


250 MS/s Dual Channel Digital Storage Oscilloscope PM3320

Operating Manual/Gebrauchsanleitung/Notice d'emploi

4822 872 00352

861215



PHILIPS

IMPORTANT

In correspondence concerning this instrument please quote the type number and serial number as given on the type plate.

NOTE: *The design of this instrument is subject to continuous development and improvement. Consequently, this instrument may incorporate minor changes in detail from the information contained in this manual.*

WICHTIG

Bei Schriftwechsel über dieses Gerät wird gebeten, die genaue Typenbezeichnung und die Gerätenummer anzugeben. Diese befinden sich auf dem Leistungsschild.

BEMERKUNG: *Die Konstruktion und Schaltung dieses Geräts wird ständig weiterentwickelt und verbessert. Deswegen kann dieses Gerät von den in dieser Anleitung stehenden Angaben abweichen.*

IMPORTANT

RECHANGE DES PIECES DETACHEES (Réparation).

Dans votre correspondance et dans vos réclamations se rapportant à cet appareil, veuillez TOUJOURS indiquer le numéro de type et le numéro de série qui sont marqués sur la plaquette de caractéristiques.

REMARQUES: *Cet appareil est l'objet de développements et améliorations continuels. En conséquence, certains détails mineurs peuvent différer des informations données dans la présente notice d'emploi et d'entretien.*

CONTENTS		Page
1.0	OPERATORS SAFETY	1-1E
1.1	Introduction	1-1E
1.2	Safety precautions	1-1E
1.3	Caution and warning statements	1-1E
1.4	Symbols	1-1E
1.5	Impaired safety-protection	1-1E
2.0	INTRODUCTION	2-1E
3.0	INSTALLATION INSTRUCTIONS	3-1E
3.1	Initial inspection	3-1E
3.2	Safety instructions	3-1E
3.2.1	Earthing	3-1E
3.2.2	Mains voltage cord and fuses	3-2E
3.3	Memory back-up batteries	3-3E
3.3.1	General information	3-3E
3.3.2	Installation of batteries	3-3E
3.4	The front cover	3-4E
3.4.1	Removing and fitting of the front-cover	3-4E
3.4.2	Access to inner-cover storage space	3-5E
3.5	Operating position of the instrument	3-6E
3.6	IEEE 488/RS232-C interface option	3-6E
4.0	OPERATING INSTRUCTIONS	4-1E
4.1	Switching-on and power-up routine	4-1E
4.1.1	Switching-on	4-1E
4.1.2	Power-up routine	4-1E
4.1.3	Default settings after switching-on	4-1E
4.2	Explanation of controls and sockets	4-2E
4.2.1	Softkeys	4-3E
4.2.2	Screen lay-out and text areas	4-7E
4.2.3	Auto-set	4-10E
4.2.4	CRT section	4-11E
4.2.5	Vertical section + menu structure	4-12E
4.2.6	Horizontal section + menu structure	4-27E
4.2.7	Trigger section + menu structure	4-30E
4.2.8	Cursor section + menu structure	4-42E
4.2.9	Memory section + menu structure	4-51E
4.2.10	Display section + menu structure	4-60E
4.2.11	Front no. section + menu structure	4-77E
4.2.12	Miscellaneous functions and remote mode	4-89E
4.2.13	Rear panel	4-91E

4.3	Detailed Operating Information	4-92E
4.3.1	Introduction	4-92E
4.3.2	Start-up procedure	4-93E
4.3.3	Use of internal registers	4-94E
4.3.4	Use of probes	4-95E
4.3.5	Input coupling AC, 0, DC	4-95E
4.3.6	Offset and shift	4-95E
4.3.7	Add mode and common-mode measurements	4-96E
4.3.8	Triggering	4-96E
4.3.9	Trigger delay	4-99E
4.3.10	Time-base modes	4-100E
4.3.11	Min/max mode	4-106E
4.3.12	Average mode	4-107E
4.3.13	Horizontal magnifier	4-108E
4.3.14	Vertical magnifier	4-108E
4.3.15	A versus B mode	4-110E
4.3.16	Analog plot mode	4-110E
4.4	Principle of operation	4-118E
4.4.1	General	4-118E
4.4.2	The signal acquisition section	4-118E
4.4.3	The storage section	4-123E
4.4.4	The display section	4-123E
4.4.5	The control section	4-125E
4.4.6	The front section	4-125E
4.4.7	The power supply section	4-128E
4.5	Brief Checking Procedure	4-126E
4.5.1	General information	4-126E
4.5.2	Preliminary settings	4-126E
4.5.3	Trace rotation	4-126E
4.5.4	Use of probes	4-127E
4.5.5	Vertical	4-127E
4.5.6	Horizontal	4-128E
4.5.7	Triggering	4-128E
4.5.8	Display	4-129E
4.5.9	Memory	4-129E
4.5.10	Cursor control	4-129E
4.5.11	Miscellaneous	4-129E
5.0	PREVENTIVE MAINTENANCE	5-1E
5.1	General information	5-1E
5.2	Removing the bezel and the contrast filter	5-1E
5.3	Replacing the memory back-up batteries	5-2E
5.4	Recalibration	5-2E

6.0	CHARACTERISTICS	6-1E
6.1	Cathode ray tube	6-2E
6.1.1	Type	6-2E
6.1.2	Usefull screen	6-2E
6.1.3	Screen type	6-2E
6.1.4	Total acceleration voltage	6-2E
6.1.5	Spot size	6-2E
6.1.6	Maximum trace distortion	6-2E
6.1.7	Graticule	6-2E
6.1.8	Orthogonality	6-2E
6.1.9	Intensity	6-2E
6.1.10	Focus	6-3E
6.1.11	Trace rotation	6-3E
6.2	Signal acquisition	6-3E
6.2.1	Sampling type	6-3E
6.2.2	Maximum sampling rate	6-3E
6.2.3	Vertical resolution	6-3E
6.2.4	Horizontal resolution	6-3E
6.2.5	Record length	6-4E
6.2.6	Acquisition time	6-4E
6.2.7	Sources	6-4E
6.2.8	Acquisition modes	6-4E
6.2.9	Maximum time difference	6-4E
6.3	Channels A and B	6-4E
6.3.1	Input connector	6-4E
6.3.2	Input impedance (in high Z position)	6-5E
6.3.3	Input impedance (in 50 ohm position)	6-5E
6.3.4	Input coupling	6-5E
6.3.5	Max. input voltage	6-5E
6.3.6	Deflection coefficient	6-5E
6.3.7	Dynamic range	6-6E
6.3.8	D.c. offset control	6-6E
6.3.9	Shift range	6-6E
6.3.10	Frequency response (in 50 ohm position)	6-6E
6.3.11	Freq. resp. (in high Z pos. through probe)	6-7E
6.3.12	Bandwidth limiter	6-7E
6.3.13	Pulse response (in 50 ohm position)	6-7E
6.3.14	Pulse resp. (in high Z pos. through probe)	6-7E
6.3.15	Max. base line instability	6-8E
6.3.16	Common mode rejection ratio	6-8E
6.3.17	MIN / MAX function	6-8E
6.3.18	Average	6-9E
6.3.19	Cross talk	6-9E
6.4	Time base	6-9E
6.4.1	Modes	6-9E
6.4.2	Time coefficients	6-9E

6.5	Trigger	6-10E
6.5.1	Sources	6-10E
6.5.2	Input connectors	6-10E
6.5.3	Input impedance of EXT trigger	6-10E
6.5.4	Coupling	6-10E
6.5.5	Max. input voltage	6-10E
6.5.6	Signal trigger sensitivity	6-11E
6.5.7	Slope selection	6-11E
6.5.8	Signal level control range	6-11E
6.5.9	Frequency response	6-11E
6.5.10	Trigger delay	6-12E
6.6	Memory	6-12E
6.6.1	Memory size	6-12E
6.6.2	Functions	6-12E
6.7	Display	6-13E
6.7.1	Sources	6-13E
6.7.2	Display expansion	6-13E
6.7.3	Display manipulations	6-13E
6.7.4	Position range	6-13E
6.8	Setting memory	6-14E
6.8.1	Memory size	6-14E
6.8.2	Functions	6-14E
6.9	Calculation facilities	6-15E
6.9.1	Functions	6-15E
6.10	Auto setting	6-15E
6.10.1	Settling time	6-15E
6.10.2	CRT functions	6-15E
6.10.3	Display functions	6-15E
6.10.4	Cursors	6-15E
6.10.5	Text	6-15E
6.10.6	Vertical acquisition	6-16E
6.10.7	Horizontal acquisition	6-17E
6.10.8	Triggering	6-17E
6.11	Cursors	6-18E
6.11.1	Cursor intensity control	6-18E
6.11.2	Horizontal resolution	6-18E
6.11.3	Vertical resolution	6-18E
6.11.4	Read out resolution	6-18E
6.11.5	Voltage cursors	6-18E
6.11.6	Time cursors	6-18E
6.12	Calibrator	6-18E
6.12.1	Wave form	6-18E
6.12.2	Internal impedance	6-18E
6.12.3	Output voltage	6-18E
6.12.4	Output current	6-18E
6.12.5	Frequency	6-18E
6.13	Power supply	6-19E
6.13.1	Source voltage	6-19E
6.13.2	Source frequency	6-19E
6.13.3	Source waveform characteristics	6-19E
6.13.4	Allowable power source interruption	6-19E
6.13.5	Power consumption	6-19E

6.14	OPTIONS	6-20E
6.15	Sundries	6-21E
6.15.1	Data and settings retention	6-21E
6.15.2	Probe read out	6-22E
6.15.3	Analog plot output	6-23E
6.16	Mechanics	6-23E
6.16.1	Height	6-23E
6.16.2	Width	6-23E
6.16.3	Depth	6-23E
6.16.4	Mass	6-23E
6.16.5	Operating position	6-23E
6.16.6	Finish	6-23E
6.16.7	Printed circuit boards	6-23E
6.16.8	Cooling	6-23E
6.17	Environmental	6-24E
6.17.1	General	6-24E
6.17.2	Meets environmental requirements of	6-24E
6.17.3	Temperature	6-24E
6.17.4	Maximum humidity	6-24E
6.17.5	Maximum altitude	6-24E
6.17.6	Vibration	6-25E
6.17.7	Shock	6-25E
6.17.8	Bench handling	6-25E
6.17.9	Salt atmosphere	6-25E
6.17.10	EMI meets requirements of	6-26E
6.17.11	Magnetic radiated susceptibility	6-26E
6.17.12	Packing	6-26E
6.17.13	Transportation	6-26E
6.18	Safety	6-26E
6.18.1	Meets requirements of	6-26E
6.18.2	Approvals	6-26E
6.19	Accessories	6-27E
6.19.1	Accessories furnished with instrument	6-27E
6.20	Optional versions	6-27E
6.20.1	General	6-27E
6.20.2	Power cord	6-27E
6.20.3	Cabinet	6-27E
6.20.4	IEEE 488/RS 232C interface	6-27E
7.0	PM3320 VERSIONS – ADDITIONAL INFORMATION	7-1E
8.0	ACCESSORY INFORMATION	8-1E
8.1	Accessories supplied with the instrument	8-1E
8.1.1	Passive probe PM8929/09 with automatic range indication	8-1E
8.1.2	Blue contrast filter	8-6E
8.1.3	Oscilloscope with front cover	8-6E
8.2	Optional accessories	8-6E
8.2.1	IEEE 488/RS232-C bus interface PM 8956	8-6E

LIST OF FIGURES

Page

Figure 2.1	250 Megasamples/s digital storage oscilloscope	2-1E
3.1	Rear view showing the fuse-holder	3-2E
3.2	Rear view showing the battery compartment	3-3E
3.3	Removing the front cover	3-4E
3.4	Removing the inner front cover (storage space)	3-5E
3.5	Carrying handle rotation and sloping positions	3-6E
4.1	Front panel function lay out	4-2E
4.2	Front panel view (softkeys)	4-3E
4.3	Example softkey text	4-4E
4.4	Screen lay-out	4-7E
4.5	Example screen lay-out	4-9E
4.6	Front panel view (auto-set)	4-10E
4.7	Front panel view (CRT section)	4-11E
4.8	Front panel view (vertical section)	4-12E
4.9	Vertical mode menu structure	4-14E
4.10	Vertical coupling menu structure	4-18E
4.11	Front panel view (horizontal section)	4-27E
4.12	Horizontal mode menu structure	4-28E
4.13	Front panel view (trigger section)	4-30E
4.14	Trigger delay menu structure	4-32E
4.15	Trigger coupling menu structure	4-36E
4.16	Trigger source menu structure	4-40E
4.17	Front panel view (cursor section)	4-42E
4.18	Cursor menu structure	4-43E
4.19	Cursor display and control	4-45E
4.20	RMS measurement	4-46E
4.21	1/dt measurement	4-47E
4.22	Mean value measurement	4-47E
4.23	Peak-peak measurement	4-49E
4.24	Rise time measurement	4-49E
4.25	Front panel view (memory section)	4-51E
4.26	Save/plot menu structure	4-53E
4.27	Front panel view (display section)	4-60E
4.28	Magnify menu structure	4-61E
4.29	Y/5 display	4-63E
4.30	Display menu structure	4-66E
4.31	Reduced text display in Y/5	4-74E
4.32	Full text display	4-76E
4.33	Front panel view (front nr. section)	4-77E
4.34	Front nr. menu structure	4-78E
4.35	Main and sub fronts	4-79E

4.36	Insert-action	4-83E
4.37	Delete-action	4-84E
4.38	Save-action	4-85E
4.39	End of sequence-action	4-88E
4.40	Front panel view (miscellaneous + remote)	4-89E
4.41	Rear panel view	4-91E
4.42	Screen of R0 contents in dual channel mode	4-93E
4.43	Display of four registers in Y/5 mode	4-94E
4.44	Suppression of common-mode signals	4-96E
4.45	Trigger filter responses	4-97E
4.46	Triggering after a number of events	4-100E
4.47	Maximum resolution after two sweeps	4-101E
4.48	Random sampling	4-103E
4.49	ROLL-mode action	4-105E
4.50	Principle of peak detection	4-106E
4.51	Principle of peak detection	4-108E
4.52	Y/5 mode	4-108E
4.53	Y*1 mode	4-109E
4.54	Y*5 mode	4-109E
4.55	Blockdiagram	4-120E
4.56	Display of trace memory contents	4-124E
5.1	Removing the bezel and contrast filter	5-1E
6.1	Input resistance $R_{par.}$ and capacitance $C_{par.}$ versus frequency	6-28E
6.2	Maximum input voltage (peak to peak) derating versus frequency	6-28E
7.1	Rear panel with indicated type plate position	7-1E
7.2	Mains connectors	7-2E
8.1	Input impedance v.s. frequency	8-4E
8.2	AC component (pk) of max. input voltage v.s. frequency	8-4E
8.3	Over-compensation	8-5E
8.4	Correct compensation	8-5E
8.5	Under-compensation	8-5E
8.6	Blue contrast filter	8-6E
8.7	Oscilloscope with front cover	8-6E

1.0 OPERATORS SAFETY

Read this page carefully before installation and use of the instrument.

1.1 INTRODUCTION

The instrument described in this manual is designed to be used by properly-trained personnel only. Adjustment, maintenance and repair of the exposed equipment shall be carried out only by qualified personnel.

1.2 SAFETY PRECAUTIONS

For the correct and safe use of this Instrument it is essential that both operating and service personnel follow generally-accepted safety procedures in addition to the safety precautions specified in this manual. Specific warning and caution statements, where they apply, will be found throughout the manual. Where necessary, the warning and caution statements and/or symbols are marked on the apparatus.

1.3 CAUTION AND WARNING STATEMENTS

CAUTION: is used to indicate correct operating or maintenance procedures in order to prevent damage to or destruction of the equipment or other property.

WARNING: calls attention to a potential danger that requires correct procedures or practices in order to prevent personnel injury.

1.4 SYMBOLS



Read the operating instructions.

1.5 IMPAIRED SAFETY PROTECTION

Whenever it is likely that safety-protection has been impaired, the instrument must be made inoperative and be secured against any unintended operation. The matter should then be referred to qualified technicians. Safety protection is likely to be impaired if, for example, the instrument fails to perform the intended measurements or shows visible damage.

2.0 INTRODUCTION

This compact dual channel digital storage oscilloscope features an extensive sampling rate of 250 Megasamples/s with a vertical bandwidth of 200 MHz and a vertical resolution of 10 bit. An outstanding feature is the AUTO-SET pushbutton facility, which automatically sets various controls of the instrument to suit the input signal value. In this way, optimum ease of operation is obtained as the input signal immediately presents a correct, stable display on the bright C.R.T. screen.

The brightness is independent of the time base settings. The M68000 microprocessor gives a wide choice of measurement and display possibilities, which can be selected via the ergonomic design- ned front panel.





Figure 2.1 250 Megasamples/s digital storage oscilloscope.

The oscilloscope is provided with integrated circuits (including thin-film circuits), which guarantee highly-stable operation.

Furthermore, connection to the local mains is simplified by a tapless switched-mode power supply that covers most voltage ranges in use:
90 V...264 V a.c.

All these features make this oscilloscope suitable for a wide range of measuring applications.



3.0 INSTALLATION INSTRUCTIONS

3.1 INITIAL INSPECTION

Check the contents of the shipment for completeness and note whether any damage has occurred during transport. If the contents are incomplete, or there is damage, a claim should be filed with the carrier immediately, and the Philips Sales or Service organisation should be notified in order to facilitate the repair or replacement of the instrument.

3.2 SAFETY INSTRUCTIONS

3.2.1 Earthing

Before any connection to the input connectors is made, the instrument shall be connected to a protective earth conductor via the three-core mains cable; the mains plug shall be inserted only into a socket outlet provided with a protective earth contact. The protective action may not be negated by the use of an extension cord without protective conductor.

WARNING: Any interruption of the protective conductor inside or outside the instrument is likely to make the instrument dangerous. Intentional interruption is prohibited.

When an instrument is brought from a cold into a warm environment, condensation may cause a hazardous condition. Therefore, make sure that the earthing requirements are strictly adhered to.

3.2.2 Mains voltage cord and fuses

Different power cords are available for the various local mains voltage outlets. The power cord version delivered is determined by the particular instrument version ordered (see also Chapter 7).

NOTE: *If the mains plug has to be adapted to the local situation, such adaption should be done only by a qualified technician.*

This oscilloscope has a power supply that covers most voltage ranges in use: 90 V...264 V a.c. (r.m.s.). This obviates the need to adapt to the local mains voltage by means of switch setting. The mains frequency range is 45 Hz...440 Hz.

WARNING: **The instrument shall be disconnected from all voltage sources when replacing a fuse.**

Mains fuse rating: 2.5 A delayed-action

The mains fuseholder is located on the rear panel (see Fig. 3.1). If the mains fuse needs replacing, proceed as follows:

- switch the instrument off and disconnect it from the mains voltage.
- remove the inner part of the fuseholder by means of a screwdriver.
- fit a new fuse of the correct rating and refit the inner part of the fuseholder.

WARNING: **Make sure that only fuses of the required current rating, and of the specified type, are used for renewal. The use of repaired fuses, and/or short-circuiting of the fuseholder, is prohibited.**



Figure 3.1 Rear view showing the fuse-holder.

3.3 MEMORY BACK-UP BATTERIES

3.3.1 General information

The memory back-up circuit has the following functions:

- after a power source interruption, or when the oscilloscope is switched off in the LOCK mode, the front-settings as well as the stored data values are saved in the internal memory.
- after the power supply is restored, the oscilloscope starts up automatically.
- if the power source is interrupted while the settings are changed by the user, it might happen that the front settings are disturbed and an automatic AUTO SET is performed after switching on the instrument again.

3.3.2 Installation of batteries

To install the batteries, the following procedure must be followed:

- remove the cover of the battery compartment located on the rear panel, by pressing the two locking tongues towards each other and pulling (see Fig. 3.2). The battery holders are now accessible.
 - insert the two penlight batteries, paying attention to the polarity indication marked on the holder (also on Fig. 3.2).
- CHECK POLARITY TO ENSURE CORRECT INSTALLATION!**
- refit the push-on cover to the rear panel.

NOTE: *It is advisable to remove the batteries when the oscilloscope is stored for longer periods than 24 hours at ambient temperatures below -30°C or above 60°C .*

IMPORTANT: *Under no circumstances should the batteries be left in the oscilloscope at ambient temperatures outside the rated range of the battery specifications!*




Figure 3.2 Rear view showing the battery compartment.

3.4 THE FRONT COVER

3.4.1 Removing and fitting of the front-cover

For ease of removal and fitting, the front cover has been designed as a simple push-fit on the front of the instrument. The front can be removed as follows:

- depress the pushbuttons in the brackets and turn the carrying handle as far as possible to the upper side of the oscilloscope.
- pull both clamping lips of the front cover outwards.
- lift the cover off the instrument.




Figure 3.3 Removing the front cover.

The cover can be refitted by simply pushing it on the oscilloscope.

3.4.2 Access to inner-cover storage space

Storage space for accessories such as probes, is available behind the inner front cover. This inner-cover can be lifted out by pressing the two locking tongues towards each other as indicated.




Figure 3.4 Removing the inner front cover. (storage space)

3.5 OPERATING POSITION OF THE INSTRUMENT

The oscilloscope may be used in the following positions:

- horizontally on its bottom feet;
- on the carrying handle in various sloping positions.

The available oscilloscope angles with respect to the working surface can be selected after depressing the push-buttons in the brackets of the carrying handle and turning. The characteristics given in Chapter III are fully guaranteed for all the above-mentioned positions.

ATTENTION: Ensure that the fan hole in the rear cover and the holes in the cover are free from obstruction, for correct functioning of the fan.

Do not position the oscilloscope on any surface which radiates heat, or in direct sunlight.




Figure 3.5 Carrying handle rotation and sloping positions.

3.6 IEEE 488/RS232-C INTERFACE OPTION PM 8956

If your oscilloscope is equipped with an option containing the IEEE 488 bus interface and an RS232-C interface, it can be used in an IEEE bus system configuration and in a system for serial communication.

For installation instructions, see information delivered with this option.

See also section 8.2 "OPTIONAL ACCESSORIES".

4.0 OPERATING INSTRUCTIONS

This chapter outlines the procedures and precautions necessary for operation. It identifies and briefly describes the functions of the front and rear panel controls and indicators, and explains the practical aspects of operation to enable an operator to evaluate quickly the instrument's main functions.

4.1 SWITCHING-ON AND POWER-UP ROUTINE

4.1.1 Switching-on



After the oscilloscope has been connected to the mains (line) voltage in accordance with Section 3.2.1. and 3.2.2. it can be switched on with the POWER ON/OFF switch on the front panel. The associated POWER indicator lamp is adjacent to the POWER ON/OFF switch.

Having switched on the oscilloscope, a power-up routine is performed after which the instrument is ready for use.

With normal installation, according to Chapter 3, and after a warming-up time of 30 minutes, the characteristics according to Chapter 6 are valid.

4.1.2 Power-up routine

When switching-on the instrument, note that the internal micro-processor automatically starts a test for a number of internal circuits.

If during this test a circuit is found to be faulty, the test stops and this will be indicated as follows:

- the instrument fails to operate normally
- some, but not all of the indicator lamps light

If this occurs, it is recommended to switch off the instrument and switch on again after a few seconds.

IMPORTANT: If the fault condition persists, contact your local PHILIPS service department.

If the system blocks during operation, it may be due to extremely high static voltages. In this event, an automatical reset of the microprocessor system is performed and the operation of the instrument is restored.

4.1.3 Default settings after switching-on.

If no back up batteries are installed and the instrument is switched on, an automatic AUTO-SET action is performed.

With back up batteries installed, the instrument settings at the moment of switching off are restored and the instrument starts up with the same setting.

4.2 EXPLANATION OF CONTROLS AND SOCKETS

The controls and sockets are listed according to their functional sections and a brief description of each is given.

The next front panel view shows the controls and sockets, and the functional lay-out of the various sections.




Figure 4.1 Front panel function lay-out.

Section:	See section:
SOFTKEYS SECTION	4.2.1
C.R.T. DISPLAY SECTION	4.2.2
AUTO SET SECTION	4.2.3
POWER ON & CRT CONTROLS SECTION	4.2.4
VERTICAL SECTION	4.2.5
HORIZONTAL SECTION	4.2.6
TRIGGER SECTION	4.2.7
CURSOR SECTION	4.2.8
MEMORY SECTION	4.2.9
DISPLAY SECTION	4.2.10
DISPLAY FILTER SECTION	4.2.10
SETTING MEMORY SECTION	4.2.11
ENTER KEYPAD SECTION	4.2.12
REMOTE SECTION	4.2.12

4.2.1 SOFTKEYS





Figure 4.2 Front panel view.

Eight softkeys are located directly at the right side of the C.R.T. screen. Different functions for these softkeys can be selected with the following eleven menu selection pushbuttons:

Menu selection pushbutton:	See section:	Menu:	See section:
VERTICAL	4.2.5	MODE	4.2.5.1
		COUPLING	4.2.5.2
HORIZONTAL	4.2.6	MODE	4.2.6.1
TRIGGER	4.2.7	DELAY	4.2.7.1
		COUPLING	4.2.7.2
		SOURCE	4.2.7.3
CURSOR CONTROL	4.2.8	CURSOR	4.2.8.1
MEMORY	4.2.9	SAVE/PLOT	4.2.9.1
DISPLAY	4.2.10	MAGNIFY	4.2.10.1
		DISPLAY	4.2.10.2
SETTING	4.2.11	FRONT No	4.2.11

After depressing one of the above mentioned pushbuttons, the selected softkey menu is displayed in the softkey text area on the C.R.T. screen, directly at the left side of the softkeys. In general the selected functions are intensified displayed in this softkey text area.

When for example the frontpanel pushbutton **VERTICAL MODE** is depressed, the following menu text will appear in the softkey text area.



When the displayed softkey function is followed by a **>**, this indicates that the relevant softkey can be used to select the next lower function menu level.



RETURN: By depressing the softkey with the function RETURN (always the last softkey in the row of 8) a jump is made to the next higher function menu level.


STOP or

EXECUTE: For some menu's a function STOP or EXECUTE is selectable which gives an automatic RETURN to the next higher function menu level after the STOP or EXECUTE action.

Figure 4.3 Example softkey text.

For the detailed explanation of the various menu structures the following explanation structure is used:





Explanation of the next lower menu level which in this example is selected by pressing softkey number three.

4	3	1	Function
4	3	2	Function
4	3	3	Function
4	3	4	Function
4	3	5	Function
4	3	6	Function
4	3	7	Function
4	3	8	RETURN

Explanation

Explanation

Explanation

Explanation

Explanation

Explanation

Explanation

Return to the next higher menu level

4	4	Function
4	5	Function
4	6	Function
4	7	Function
4	8	RETURN

Explanation

Explanation

Explanation


Explanation

Return to the highest menu level

5	Function
6	Function
7	Function
8	Function

Example:

If via the VERTICAL MODE menu the factor C=2 has to be selected for the AVERAGE mode, it can be done in the following way:



- Press frontpanel pushbutton VERTICAL MODE
- Press softkey 7 function PROCESSING>
- Press softkey 4 function AVERAGE>
- Press softkey 2 function C=2



4.2.2**Screen lay-out and text areas**

Figure 4.4 Screen lay-out.


Text areas:

ACTUAL FRONT SETTINGS AND CURSOR CALC. RESULTS IN THE TOP TEXT AREA




REDUCED TEXT IN THE TRACE AREA

(selectable via the DISPLAY menu if Y/5 is selected via the MAGNIFY menu)




FULL TEXT IN THE BOTTOM TEXT AREA

(selectable via the DISPLAY menu)



FRONT TEXT IN THE BOTTOM TEXT AREA

(selectable via the DISPLAY menu)



General information about screen lay out and texts:

- Activated functions are always intensified displayed.
- Only softkey functions that can be activated in the actual oscilloscope setting, are displayed.
- Only the registers that are selected for display, are visible in the softkey text area.
- A letter A and/or B can be displayed near the start of each trace on the screen for channel identification. (selectable via the DISPLAY menu).
- If AC or DC triggering is selected, a mark \uparrow , \downarrow or \times is displayed on the left side of the screen for trigger level indication. For LINE triggering this is a mark L and for EXT triggering this is a mark X.
- During the selection of the trigger level for EVENTS, a letter E is displayed on the left side of the screen for trigger level indication.
- \sim indicates lower accuracy for this function; is also indicated by pilot lamps
- / indicates positive trigger slope selection
- \ indicates negative trigger slope selection
- ? indicates that the value can not be calculated in the selected mode
- \diamond indicates MIN / MAX mode selection
- dts indicates that a number of dots is displayed

Example:




Figure 4.5 Example screen lay-out.

6471023

4.2.3 AUTO-SET




Figure 4.6 Front panel view.



If AUTO is pushed, a number of functions are preselected. Some of these functions are selected in relation to the applied input signals.

During the AUTO SET action, a message

***** AUTO SET BUSY *****

is displayed.

Settings are such, that the contents of register R0 are displayed over the full C.B.T. screen.

The input signals will be displayed as triggered signals of a few periods and with an amplitude of a few divisions.

Details about AUTO SETTING are given in section 6.10 of the CHARACTERISTICS.

In LOCK mode ■ message

No AUTO SET possible in LOCK mode.

is displayed.

4.2.4. CRT section




Figure 4.7 Front panel view.



POWER ON/OFF switch

After switching on the instrument, first an automatic power-up test is started (see also section 4.1).
The POWER ON LED indicates when the instrument is switched-on.



Continuously-variable control of the graticule illumination.



Screwdriver control for aligning the trace in parallel with the horizontal graticule lines.



Continuously-variable control of the trace brilliance on the C.R.T. screen.



Screwdriver control for the focussing of the C.R.T. electron beam (including the C.R.T. text).



Continuously-variable control of the brilliance of the C.R.T. text (control settings, cursors, softkey functions and messages).

4.2.5. VERTICAL SECTION AND MENU STRUCTURE




Figure 4.8. Front panel view.



This UP/DOWN control permits selection of the vertical deflection coefficients of the relevant channel in 10 steps from 5 mV/div..5 V/div in a 1-2-5 sequence.

The selected vertical deflection coefficient for the relevant channel is displayed in the top text area of the C.R.T. screen.

If the left side of the UP/DOWN control (V) is pushed, the amplitude of the signal on the screen decreases. This means that the AMPL/DIV value becomes bigger (e.g. the attenuator jumps from 2 V/div to 5 V/div).

If the right side of the UP/DOWN control (mV) is pushed, the amplitude of the signal on the screen increases. This means that the AMPL/DIV value becomes smaller (e.g. the attenuator jumps for 5 V/div to 2 V/div).



Continuously variable control of the deflection coefficients of the relevant channel. The deflection coefficients of the relevant channel are only calibrated when this control is set fully clockwise (UNCAL LED is off). In the UNCAL position, the vertical sensitivity decreases. The complete sensitivity range between two AMPL/DIV settings is adjustable.

The adjusting speed increases after turning continuously in one direction. After stopping, starting in the reverse direction resumes with a slow adjusting speed.



Pilot lamp indicating lower accuracy for the relevant vertical deflection coefficients.

If the VARIABLE control of the relevant channel is turned fully clockwise, the UNCAL pilot lamp is switched off.



Continuously variable control giving vertical input shift of the trace of the relevant channel.

This shift is introduced behind the attenuator but before the memory.

The adjusting speed increases after turning continuously in one direction. After stopping, starting in the reverse direction resumes with a slow adjusting speed.



BNC input socket for the channel with probe indication detector.



BNC output socket for a square-wave calibration signal with an amplitude of 1 Vp-p and a frequency of 2 kHz.

The output voltage is reduced by a half when the output is terminated in 50 ohm.

When the output is short-circuited, the output current is 20 mA p-p (reduced by a half when terminated in 50 ohm).

The zero line of the square-wave signal is on the base-line level.



If pushbutton MODE is pushed, the VERTICAL MODE menu is displayed. See 4.2.5.1.



If pushbutton COUPLING is pushed, the VERTICAL COUPLING menu is displayed. See 4.2.5.2.

4.2.5.1. VERTICAL MODE MENU STRUCTURE





Figure 4.9 Vertical mode menu structure.

VERTICAL MODE MENU



After selection of the VERTICAL MODE menu by depressing pushbutton VERTICAL MODE, various modes related to the signal acquisition can be selected.

1 A

Pushing softkey A gives signal input from channel A only, indicated by an intensified text A.

If softkey A is pushed and function A and B was operative, in A versus B - mode a message will be displayed with the text

**No A versus B possible with one channel only.
A versus B is switched off.**

and channel A will be selected.

2 A and B

Pushing softkey A and B gives signal input from both channels A and B, indicated by an intensified text A and B.

3 ADD

Pushing softkey ADD gives signal input from channel A and channel B and the algebraic sum of A and B is recorded in the memory, indicated by an intensified text ADD. If softkey ADD is pushed and A and B was operative, in A versus B - mode, a message will be displayed with the text

**No A versus B possible with one channel only.
A versus B is switched off.**

and ADD will be selected.

4 B

Pushing softkey B gives signal input from channel B only, indicated by an intensified text B.

If softkey B is pushed and A and B was operative, in A versus B - mode a message will be displayed with the text

**No A versus B possible with one channel only.
A versus B is switched off.**


and channel B will be selected.

5 A-INVERT

6 B-INVERT

After pushing the relevant INVERT softkey the signal of the relevant channel is inverted before it is digitized and recorded in the memory. The text A-INVERT or B-INVERT is then intensified displayed.

7 PROCESSING>



After selecting PROCESSING>, the VERTICAL PROCESSING menu is displayed and input signal processing can be selected.

The text MIN / MAX is not visible in the time base positions faster than 5 us/div.

The text AVERG OFF is only visible if average is selected.

7 1 MIN / MAX

After pushing softkey MIN / MAX, two peak detectors are operative for each channel.

Glitches in the A channel as well as in the B channel can be caught at the same moment, since samples for both channels are taken at the same moment.

When switching from 5 us/div to 2 us/div with MIN / MAX on the MIN / MAX limitation is indicated by a message

No MIN / MAX possible when time base faster than 5 us/div. MIN / MAX switched off.

When switching from 2 us/div to 5 us/div and MIN / MAX on, the following message is displayed

Previous MIN / MAX setting restored.

- Glitches longer than 3 ns are detected with an accuracy of more than 50 %.
- Glitches shorter than 20 ns are not detected if occurring during the reset time of the peak detectors.
- MIN / MAX is not possible in combination with the AVERAGE function. If MIN / MAX is selected when AVERAGE is active, a message

**MIN / MAX and AVERAGE not possible.
MIN / MAX on, AVERAGE off.**

is displayed.


7 2 BW LIMIT

After pushing softkey BW LIMIT a bandwidth filter is active. This is indicated by an intensified text BW LIMIT.

The bandwidth is 3 dB down at 20 MHz, using 3 dB/octave filter. The filter is active in both channels.

7 3 AVERG OFF

After pushing softkey AVERG OFF, AVERAGE is switched off. This function is only displayed when AVERAGE is selected.



After selecting AVERAGE, the VERTICAL AVERAGE menu is displayed and the average function can be selected. (See also section 4.3.12).

As long as average is operative, this is indicated by an intensified text C= in the VERTICAL PROCESSING menu.

If AVERAGE is not operative, it is indicated by an intensified text C=OFF in the VERTICAL PROCESSING menu.

AVERAGE is not possible in combination with the MIN / MAX function.

If AVERAGE is selected when MIN / MAX is active, a message

AVERAGE and MIN / MAX not possible.

AVERAGE on, MIN / MAX off.

is displayed.

Softkey OFF must be pushed to switch the AVERAGE function off.

7 4 2 C=2

7 4 3 C=4

7 4 4 C=8

7 4 5 C=16

7 4 6 C=32

7 4 7 C=64

Different calculation constants between C=2 and C=64 can be selected. The bigger the value of C is, the stronger the AVERAGE effect is. For more detailed information see 4.3.12.

7 4 8 RETURN

After pushing softkey RETURN, the VERTICAL PROCESSING menu is displayed again and the average selection (the value of C) remains as selected before.

7 5 —

7 6 —

7 7 —

7 8 RETURN

After pushing RETURN, the VERTICAL MODE menu is displayed again and the selections previously selected remain.

8 —

4.2.5.2. VERTICAL COUPLING MENU STRUCTURE



VERTICAL
COUPLING

Figure 4.10 Vertical coupling menu structure.

VERTICAL COUPLING MENU



After pressing pushbutton VERTICAL COUPLING, the VERTICAL COUPLING menu for channel A and channel B is displayed and the input coupling for these channels can be selected.

1 AC DC (channel A)

With this softkey a selection can be made between AC input coupling and DC input coupling. Pushing this softkey changes the selection from AC to DC or reverse, the active selection is displayed intensified.

If AC is selected, the offset becomes 0 Volt.

If DC is selected, the previously selected offset value is active again.

2 GROUND (channel A)

With GROUND selected, the connection between the input socket and the relevant input circuit is interrupted and the input circuit is grounded.

3 1 MOhm 50 Ohm (channel A)

With this softkey, a selection can be made between an input impedance of 1 MOhm or an input impedance of 50 Ohm.

If a probe with automatic probe indication is used, the oscilloscope automatically switches to the correct impedance. No other selection can be made then.

If an input voltage higher than 5 V dc and ac r.m.s. is applied to the input socket and 50 Ohm is selected, the instrument does not switch to 50 Ohm and a message

**Unsafe condition for 50 Ohm to be switched on.
Input impedance not changed.**

is displayed.

If 50 Ohm is on and an input voltage between 5 V dc and ac r.m.s. and 25 V dc and ac r.m.s. is applied to the input socket, the instrument will automatically switch to 1 MOhm input impedance (without warning).


If 50 Ohm is on and an input voltage higher than 25 V dc and ac r.m.s. is applied to the input socket; 50 Ohm input impedance will *not* be switched off and the message

OVERLOAD ON CHANNEL A: reduce signal to prevent damages.

is displayed.

IMPORTANT: Remove the input voltage or decrease the input voltage immediately to ≤ 5 V dc and ac r.m.s.!

4 OFFSET> (channel A)



If OFFSET is selected the VERTICAL/OFFSET A menu is displayed and an offset (vertical shift before attenuator) value can be selected.

OFFSET is only effective if DC input coupling is selected.

If OFFSET is selected and AC coupling is active, a message

**OFFSET only possible if channel is DC- coupled.
AC is switched to DC.**

is displayed and the input coupling is automatically switched to DC.

4 1 AUTO

With softkey AUTO, the offset value is set to such a level, that the mid-value of the input signal is shifted as much as possible to mid-memory.

A message

*** AUTO OFFSET FOR CHANNEL A ***

is displayed.

The new offset value is displayed in the softkey text area. The VERTICAL SHIFT is set to zero.

A message

AUTO OFFSET error : signal offset out of range.


is displayed if the input signal is exceeding the offset control range. In this case the original offset value is restored.

4 2 —

4 3 ZERO

With softkey ZERO, the offset value is set to zero.

4 4 ENTER>



After selecting ENTER, the ENTER menu will be displayed and the offset can be selected with the numeric keypad.

The range in Volts, which is related to the actual AMPL/DIV setting, is indicated and the selected value is rounded off and made visible in the softkey text area.

A message

Too many digits: total entry is cleared.

can be displayed if too many digits are entered on the numeric keyboard.

4 4 1 --

4 4 2 --

4 4 3 --

4 4 4 --

4 4 5 --

4 4 6 CLEAR

If an error is made, the selected offset value can be cleared after pushing softkey CLEAR.

4 4 7 --

4 4 EXECUTE

After pushing this softkey, the selected offset value is entered and an AUTO RETURN is performed to menu VERTICAL OFFSET A.

If EXECUTE is pressed after CLEAR the previous value remains (in the VERTICAL OFFSET A menu).

A message

A- OFFSET out of range.

can be displayed if the entered value exceeds the given range.

4 5 UP

Pushing softkey UP gives a more positive or less negative offset to the signal. The amount depends on the attenuator setting and the setting SCREEN or DIV. If DIV is selected (DIV intensified) one division is added to the existing offset after each push of the UP softkey.

If SCREEN is selected (SCREEN intensified) a full screen of 10 divisions is added to the existing offset after each push of the UP softkey.

The result is displayed in volts, but the voltage is recalculated from the number of divisions.

For example: when 1.005 V is visible, it becomes 1.025 V, then 1.045 V etc. if the attenuator has been set to 20 mV/div. (200 mV for a full screen).

4 6 DOWN

Pushing softkey DOWN gives a less positive or more negative offset to the signal. The amount depends on the attenuator setting and the setting SCREEN or DIV. If DIV is selected (DIV intensified) one division is subtracted from the existing offset after each push of the DOWN softkey.

If SCREEN is selected (SCREEN intensified) a full screen of 10 divisions is subtracted from the existing offset after each push of the DOWN softkey.

The result is displayed in volts, but the voltage is recalculated from the number of divisions.

For example: when 1.005 V is visible, it becomes 985 mV, then 965 mV etc. if the attenuator has been set to 20 mV/div. (200 mV for a full screen).

4 7 SCREEN DIV

With this softkey, a selection can be made between an offset change in divisions or in screens (10 divisions) when the softkeys UP or DOWN are operated.

Pushing this softkey changes the selection from SCREEN to DIV or reverse. The active selection is displayed intensified.

4 8 RETURN

After pushing softkey RETURN, menu VERTICAL COUPLING A is displayed again.

The selections as made before remain.

5 AC DC (channel B)

With this softkey a selection can be made between AC input coupling and DC input coupling. Pushing this softkey changes the selection from AC to DC or reverse, the active selection is displayed intensified.

If AC is selected, the offset becomes 0 Volt.

If DC is selected, the previously selected offset value is active again.

6 GROUND (channel B)

With GROUND selected, the connection between the input socket and the relevant input circuit is interrupted and the input circuit is grounded.

7 1 MOhm 50 Ohm (channel B)

With this softkey, a selection can be made between an input impedance of 1 MOhm or an input impedance of 50 Ohm.

If a probe with automatic probe indication is used, the oscilloscope automatically switches to the correct impedance. No other selection can be made then.

If an input voltage higher than 5 V dc and ac r.m.s. is applied to the input socket and 50 Ohm is selected, the instrument does not switch to 50 Ohm and a message

**Unsafe condition for 50 Ohm to be switched on.
Input impedance not changed.**

is displayed.


If 50 Ohm is on and an input voltage between 5 V dc and ac r.m.s. and 25 V dc and ac r.m.s. is applied to the input socket, the instrument will automatically switch to 1 MOhm input impedance (without warning).

If 50 Ohm is on and an input voltage higher than 25 V dc and ac r.m.s. is applied to the input socket; 50 Ohm input impedance will *not* be switched off and the message

OVERLOAD ON CHANNEL B: reduce signal to prevent damages.

is displayed.

IMPORTANT: Remove the input voltage or decrease the input voltage immediately to ≤ 5 V dc and ac r.m.s.]

8 **OFFSET** > (channel B)

If **OFFSET** is selected the **VERTICAL/OFFSET B** menu is displayed and an offset (vertical shift before attenuator) value can be selected.

OFFSET is only effective if **DC** input coupling is selected.

If **OFFSET** is selected and **AC** coupling is active, a message

**OFFSET only possible if channel B DC- coupled.
AC is switched to DC.**

is displayed and the input coupling is automatically switched to **DC**.

8 1 **AUTO**

With softkey **AUTO**, the offset value is set to such a level, that the mid-value of the input signal is shifted as much as possible to mid-memory.

A message

***** AUTO OFFSET FOR CHANNEL B *****

is displayed.

The new offset value is displayed in the softkey text area. The **VERTICAL SHIFT** is set to zero.

A message


AUTO OFFSET error : signal offset out of range.

is displayed if the input signal is exceeding the offset control range. In this case the original offset value is restored.

8 2 **--**8 3 **ZERO**

With softkey **ZERO**, the offset value is set to zero.

8 4 ENTER>



After selecting ENTER, the ENTER menu will be displayed and the offset can be selected with the numeric keypad.

The range, which is related to the actual AMPL/DIV setting, is indicated and the selected value is rounded off and made visible in the softkey text area.

A message

Too many digits: total entry is cleared.

can be displayed if too many digits are entered on the numeric keyboard.

8 4 1 ---

8 4 2 ---

8 4 3 ---

8 4 4 ---

8 4 5 ---

8 4 6 CLEAR

If an error is made, the selected offset value can be cleared after pushing softkey CLEAR.

8 4 7 ---

8 4 8 EXECUTE

After pushing this softkey, the selected offset value is entered and an AUTO RETURN is performed to menu VERTICAL OFFSET B.

If EXECUTE is pressed after CLEAR the previous value remains (in the VERTICAL OFFSET B menu).

A message

B- OFFSET out of range.

can be displayed if the entered value exceeds the given range.

8 5 UP

Pushing softkey UP gives a more positive or less negative offset to the signal. The amount depends on the attenuator setting and the setting SCREEN or DIV. If DIV is selected (DIV intensified) one division is added to the existing offset after each push of the UP softkey.

If SCREEN is selected (SCREEN intensified) a full screen of 10 divisions is added to the existing offset after each push of the UP softkey.

The result is displayed in volts, but the voltage is recalculated from the number of divisions.

For example: when 1.005 V is visible, it becomes 1.025 V, then 1.045 V etc. if the attenuator has been set to 20 mV/div (200 mV for a full screen).

8 6 DOWN

Pushing softkey DOWN gives a less positive or more negative offset to the signal. The amount depends on the attenuator setting and the setting SCREEN or DIV. If DIV is selected (DIV intensified) one division is subtracted from the existing offset after each push of the DOWN softkey.

If SCREEN is selected (SCREEN intensified) a full screen of 10 divisions is subtracted from the existing offset after each push of the DOWN softkey.

The result is displayed in volts, but the voltage is recalculated from the number of divisions.

For example: when 1.005 V is visible, it becomes 985 mV, then 965 mV etc. if the attenuator has been set to 20 mV/div. (200 mV for a full screen).

8 7 SCREEN DIV

With this softkey, a selection can be made between an offset change in divisions or in screens (10 divisions) when the softkeys UP or DOWN are operated.

Pushing this softkey changes the selection from SCREEN to DIV or reverse. The active selection is displayed intensified.

8 8 RETURN

After pushing softkey RETURN, menu VERTICAL COUPLING B is displayed again.

The selections as made before remain.

4.2.6. HORIZONTAL SECTION AND MENU STRUCTURE




Figure 4.11 Front panel view.



This switch permits selection of the horizontal deflection coefficients of the time base in 28 steps from 5 ns/div ... 5 s/div. In a 1-2-5 sequence. In ROLL mode from 50 ms/div ... 360 s/div.

The selected horizontal deflection coefficient is displayed in the upper text area of the C.R.T. screen.

If the left side of the UP/DOWN control(s) is pushed more signal periods will be displayed. This means that the TIME/DIV value becomes bigger (e.g. the time-base jumps from 2 ms/div to 5 ms/div). If the right side of the UP/DOWN control (ns) is pushed fewer signal periods will be displayed.

This means that the TIME/DIV value becomes smaller (e.g. the time-base jumps from 5 ms/div to 2 ms/div).

☐ ALIASED

Pilot lamp indicating that aliasing is detected on the trigger channel. If the pilot lamp is on, the wave form which is displayed on the screen does not represent the real shape of the input signal. See also 4.3.11.

☐ RSP ONLY

Pilot lamp indicating that the digital time-base is set in the random sampling mode (100 ns/div ... 5 ns/div). In this case only signals with a repetitive character can be measured.

☐ NOT TRIG'D

Pilot lamp indicating that the oscilloscope is not triggered.



If pushbutton HORIZONTAL MODE is pushed, the HORIZONTAL MODE menu is displayed. See 4.2.6.1.

4.2.6.1. HORIZONTAL MODE MENU STRUCTURE





Figure 4.12 Horizontal mode menu structure.

HORIZONTAL MODE MENU



After selection of the HORIZONTAL MODE menu by depressing pushbutton HORIZONTAL MODE, the required time-base mode can be selected.

The text MAX RESOL. is not visible in the MIN / MAX and in the AVERAGE mode.

The texts RUN STOP and TRIG.STOP are only visible if ROLL is selected.

1 RECURRENT

If the RECURRENT-mode is selected, a new signal is recorded in register R0 and afterwards after a hold off refreshed each time that the trigger level is passed and that the selected delay has been reached.

2 MULT. ARMD

In MULTIPLE-mode (text MULT. ARMD intensified), four single shot signals can be captured sequentially. The signals are recorded in R3, R2, R1 and the last one in R0. The end of the cycle is indicated by a low intensified text ARMD.

A new cycle can be started by pushing softkey MULT. ARMD or pushbutton CLEAR and the text ARMD is intensified until the last single shot signal is recorded.

3 SING. ARMD

In SINGLE-mode (text SING. ARMD intensified) the contents of register R0 will be overwritten by a new signal when the trigger level is passed and the selected delay has been reached.

This is indicated by a low intensified text ARMD.

If afterwards a new single shot signal should be captured the same SING. ARMD softkey or pushbutton CLEAR should be pushed and the text ARMD is intensified until the single shot signal is recorded.

4 MAX RESOL.

In the time base range 200 ns/div. ... 500 us/div. the resolution of the display can be increased. Pushing MAX RESOL. activates a higher resolution. The pilot lamp REP ONLY indicates that a repetitive signal is required for this mode. If AVERAGE is selected, the instrument automatically selects MAX RESOL. When MIN / MAX is active the MAX RESOL. mode is deselected. In both situations the function is not displayed in the menu.

5 ROLL

If softkey ROLL is pushed, the ROLL-mode is selected and time-base speeds from 50 ms/div. ... 360 s/div. can be selected. The ROLL-mode is then automatically active. The text RUN is intensified during an active ROLL-mode operation, the input signal is then recorded in the register R0 and is moving over the C.R.T. screen from the right to the left. During the ROLL-mode the contents of register R0 can be copied via the SAVE/PLOT menu in one of the registers R1, R2 or R3 if required.

6 RUN STOP

This text is only displayed if the ROLL-mode is selected. With this softkey the ROLL-mode action can immediately be stopped or started. The selected function is intensified. In this mode no trigger selections are possible and a message

NO TRIGGER selections possible in manual ROLL mode.

is displayed if a trigger selection is done.

7 TRIG.STOP

This text is only displayed if the ROLL-MODE is selected. With this softkey the ROLL-mode action can be stopped on receipt of a trigger

If this softkey is pushed the text TRIG.STOP is intensified.

If a trigger delay zero is selected, then the ROLL mode continues until the trigger stop point is at the left side of the screen.

Then the action stops. This is indicated by an intensified text STOP.

By selecting another trigger delay this point can be set more to the left (positive trigger delay) or more to the right (negative trigger delay = pretrigger).

A new picture can be built up by depressing softkey RUN or pushbutton CLEAR. For more details see section 4.3.10.3 (detailed operating information).

4.2.7 TRIGGER SECTION AND MENU STRUCTURE




Figure 4.13 Front panel view.



Selection of an external applied sampling clock.

This means that the conversion rate depends on the signal which is applied to the EVENTS/EXT CLOCK input BNC socket and that the internal time-base circuit is switched off. The signal level on which the system reacts can be softkey selected via function EV/EXT CLK in menu TRIGGER COUPLING.

If EXT CLOCK is selected by pushing pushbutton EXT CLOCK, and EVENTS trigger delay was selected before, a message

**EXTERNAL CLOCK selected with EVENTS active
EVENTS switched off.**

is displayed.



EXT CLOCK

BNC input socket for external applied clock signals.
The maximum clock frequency is 50 kHz.
Pushbutton EXT CLOCK must be pressed; built in pilot lamp lights.

EVENTS

BNC input socket for external applied events for trigger delay. For this function pushbutton EXT CLOCK must be switched off.

Measuring earth socket



BNC input socket for external triggering, to be used in combination with the trigger source selection facility.



With **NEGATIVE SLOPE** off, the time base is triggered on the positive-going edge of the trigger signal. With **NEGATIVE SLOPE** pressed (on), the time base is triggered on the negative-going edge of the trigger signal, which is indicated by the built in pilot lamp. The pilot lamp lights up if **NEGATIVE SLOPE** is selected.

The adjusting speed increases after turning continuously in one direction. After stopping, starting in the reverse direction resumes with a slow adjusting speed.



Continuously-variable control to determine the **LEVEL** of the trigger point on the trigger signal at which the signal acquisition starts.



If pushbutton **DELAY** is pushed, the **TRIGGER DELAY** menu is displayed. See 4.2.7.1.



If pushbutton **COUPLING** is pushed, the **TRIGGER COUPLING** menu is displayed. See 4.2.7.2.



If pushbutton **SOURCE** is pushed, the **TRIGGER SOURCE** menu is displayed. See 4.2.7.3.

4.2.7.1 TRIGGER DELAY MENU STRUCTURE





Figure 4.14 Trigger delay menu structure.

TRIGGER DELAY MENU



After pushing the DELAY pushbutton, the TRIGGER DELAY menu is displayed and the trigger delay can be selected.

1 TIME DIV

The trigger delay can be set in seconds (TIME) or in divisions (DIV), the selected text is intensified.

When TIME is intensified, the delay is displayed in seconds, and when DIV is intensified, the delay is displayed in divisions.

2 EVENTS

The trigger delay can be set in events (EVT) after pushing softkey EVENTS, the text EVENTS is then intensified and the delay is displayed in events.

3 ZERO

Pushing softkey ZERO sets the active trigger delay value for TIME or DIV to zero. The number of EVENTS becomes 1.

This effects only the active delay: TIME DIV or EVENTS.

4 ENTER

TRIGGER DELAY	
Range:	1
-10..9999	2
Resolution	3
1 DIV	4
? DIV	5
CLEAR	6
EXECUTE	7
	8

After selecting ENTER, the TRIGGER DELAY menu will be displayed and the trigger delay can be selected with the numeric keyboard. The actual range is indicated and the value is visible in the softkey text area.

Possible texts for:

TIME : .. s (ms and us)
 DIVISIONS : .. DIV
 EVENTS : .. EVT

A message

Too many digits: total entry is cleared.

can be displayed if too many digits are entered via the numeric keyboard.

For DIV and EVENTS a message

No decimal point allowed in this enter menu.

can be displayed if a decimal point is entered via the numeric keyboard.

4	1	--
4	2	--
4	3	--
4	4	--
4	5	--
4	6	CLEAR

If an error is made, the trigger delay value can be cleared by pushing softkey CLEAR.

4	7	--
---	---	----

4 8 EXECUTE

After pushing this softkey, the selected trigger delay value is entered and an AUTO RETURN is performed to menu TRIGGER DELAY.
If after CLEAR the softkey EXECUTE is pressed the trigger delay value keeps its previous value in the TRIGGER DELAY menu.

For TIME and DIV a message

TRIGGER DELAY number out of range.

can be displayed if the entered value exceeds the given range.

For EVENTS a message

EVENT number out of range.

can be displayed if the entered value exceeds the given range.

5 UP

Pushing softkey UP increments the trigger delay.

The TIME increases in steps of the time-base setting multiplied by 1 or by 10 depending on the setting SCREEN or DIV.

DIV increases in steps of 1 division and SCREEN increases in steps of 1 screen. EVENTS always increases in steps of 1 event.

6 DOWN

Pushing softkey DOWN decrements the trigger delay. The TIME decreases in steps of the time-base setting multiplied by 1 or by 10 depending on the setting SCREEN or DIV. DIV decreases in steps of 1 division and SCREEN decreases in steps of 1 screen. EVENTS always decreases in steps of 1 event.

7 SCREEN DIV

The text SCREEN DIV is displayed when TIME or DIV is active.

If the text SCREEN is intensified, the UP and DOWN softkeys operate with 1 screen.

If the text DIV is intensified, the UP and DOWN softkeys operate with 1 division.

Pushing this softkey changes the selection from SCREEN to DIV or reverse.

If EVENTS triggering is operative, this text is not displayed.

8 —

4.2.7.2 TRIGGER COUPLING MENU STRUCTURE

TRIGGER
COUPLING





Figure 4.15 Trigger coupling menu structure.

TRIGGER COUPLING MENU



After pushing the COUPLING pushbutton, the TRIGGER COUPLING menu is displayed and different trigger modes and slopes can be selected.

The following combinations can be selected:

AC
 AC HF reject
 AC LF reject
 DC
 DC HF reject
 AUTO
 AUTO HF reject
 AUTO LF reject
 TVF (No DUAL possible)

1 AC DC

If the text AC is intensified, AC trigger coupling is selected. Pushing the softkey again will then give DC trigger coupling and visa versa.

If neither AC nor DC is intensified AUTO- or TVF-triggering is active. In this case the current trigger level is indicated by a mark I on the left side of the C.R.T. screen.

2 AUTO

If the softkey AUTO has been pushed, automatic triggering is active, which is indicated by an intensified text AUTO.

This means that the trigger level setting is limited between the highest and the lowest level of the signal. The level setting is still possible between these limits. Trigger input is AC coupled.

3 TVF


If TVF has been selected, television frame signal synchronisation is obtained (for CCIR system 625/525 lines).

Check the correct setting of the trigger slope (in accordance to the T.V. system under test). If the LEVEL control is operated, a message

Leveling not possible when in TV- mode.

is displayed.

4 EV/EXT CLK



After selecting EV/EXT CLK the COUPLING EV/EXT CLK menu is displayed and the trigger levels and trigger slope can be selected.

4 1 TTL

After pushing softkey TTL, the text TTL is intensified and the trigger level of the EVENTS/EXT CLOCK input is set for TTL signals. The trigger level is indicated by a mark \equiv on the left side of the C.R.T. screen.

4 2 ECL

After pushing softkey ECL, the text ECL is intensified and the trigger level of the EVENTS/EXT CLOCK input is set for ECL signals. The trigger level is indicated by a mark \equiv on the left side of the C.R.T. screen.

4 3 VAR

After pushing softkey VAR, the text VAR is intensified and the trigger level of the EVENTS/EXT CLOCK input can manually be set by the trigger level control. The trigger level is indicated by a mark \equiv on the left side of the C.R.T. screen and the actual value is displayed in the softkey text area.

4 4 --

4 5 --

4 6 --

4 7 SLOPE + -

Detection of events or external clock pulses on either the positive or the negative slope of the input signal on the EV/EXT CLK input can be selected. The selected function is displayed intensified.

Pushing the softkey changes the selection from + to - or reverse. This is also indicated by the pilot lamp in the pushbutton NEGATIVE SLOPE.

4 ■ RETURN

After pushing softkey RETURN, menu TRIGGER COUPLING is displayed again. The selections as made before remain and the trigger level potentiometer is operative.

In the TRIGGER COUPLING menu, the pilot lamp of the pushbutton NEGATIVE SLOPE now indicates the trigger slope.

5 LF REJECT

With LF REJECT selected, low frequency trigger signals (up to 50 kHz) are blocked. Only possible if the trigger input is AC coupled.

6 HF REJECT

With HF REJECT selected, high frequency trigger signals (> 50 kHz) are blocked.

7 SLOPE + -

With this softkey, triggering on the positive or on the negative slope of the signal can be selected. SLOPE + - overrules DUAL selection.

8 DUAL

After selection of DUAL triggering, the instrument is able to trigger on either the positive or the negative slope of the signal. The level (positive and negative) can be adjusted by the level potentiometer.

This DUAL function is specially usefull in the SINGLE-shot mode.

DUAL can not be selected in the time base range 100 ns/div ... 5 ns/div and the text DUAL is not displayed then. If DUAL is selected and this time base range is entered, a message

DUAL slope not possible in SAMPLING.

DUAL switched off.

is displayed.

4.2.7.3 TRIGGER SOURCE MENU STRUCTURE





Figure 4.16 Trigger source menu structure.

TRIGGER SOURCE MENU



After pushing the SOURCE pushbutton, the TRIGGER SOURCE menu is displayed and different trigger sources, slopes and modes can be selected.

1 A

With A selected, triggering is achieved on a signal that is internally derived from channel A.

2 B

With B selected, triggering is achieved on a signal that is internally derived from channel B.

3 EXT

With EXT selected, triggering is obtained from an external signal via the EXT TRIG input socket.

4 EXT:10

With EXT:10 selected, external triggering is obtained as above, but the signal is attenuated by a factor of 10.

5 LINE

With LINE selected, triggering is obtained by a signal internally derived from the line voltage.

6 --

7 SLOPE + -

With this softkey, triggering on the positive or on the negative slope of the signal can be selected. SLOPE + - overrules DUAL selection.

8 DUAL

After selection of DUAL triggering, the instrument is able to trigger on either the positive or the negative slope of the signal. The level (positive and negative) can be adjusted by the level potentiometer.

This DUAL function is specially useful in the SINGLE-shot mode.

DUAL can not be selected in the time base range 100 ns/div ... 5 ns/div and the text DUAL is not displayed then. If DUAL is selected and this time base range is entered, a message

**DUAL slope not possible in SAMPLING.
DUAL switched off.**

is displayed.

4.2.8 CURSOR SECTION




Figure 4.17 Front panel view.



This control functions when cursor operation is switched on via the CURSORS SELECT menu and one of the registers is selected (Rn A(B)) - (Rn ADD) or (Rn MUL) intensified).

Continuously variable control to determine the horizontal position of the 1st cursor (always the most left cursor) on the screen.

Range lies between the most left and most right side of the trace visible on the C.R.T. screen.

The adjusting speed increases after turning continuously in one direction. After stopping, starting in the reverse direction resumes with a slow adjusting speed.



This control functions when cursor operation is switched on via the CURSORS SELECT menu and one of the registers is selected (Rn A(B)) - (Rn ADD) or (Rn MUL) intensified).

Continuously variable control to determine the horizontal position of the 2nd cursor (most right cursor) on the screen.

Range lies between the most left and most right side of the trace visible on the C.R.T. screen.

The adjusting speed increases after turning continuously in one direction. After stopping, starting in the reverse direction resumes with a slow adjusting speed.



If pushbutton CURSOR is pushed, the CURSORS SELECT menu is displayed. See 4.2.8.1.

dt ...

Calculation results are displayed in the top text area of the C.R.T. screen.

dV ...

time between the cursors.

dV ... voltage difference between the cursors.

4.2.8.1 CURSOR MENU STRUCTURE





Figure 4.18 Cursor menu structure.

CURSOR MENU



After pushing pushbutton CURSOR, the CURSORS SELECT menu is displayed and the two cursors 1st and 2nd can be set at choice for various measurements using the controls 1st and 2nd.

The cursors are displayed on the previous selected positions.

As soon as the cursors (crossed lines with a length of 2 divisions) are visible, the horizontal distance between the cursors is measured and displayed in time (dt) in the top text area of the C.R.T. screen.

The vertical distance between the cursors is measured and displayed in voltage (dV).

When positioning cursor 1st and cursor 2nd, they can not pass each other. Cursor 1st is always the left one and cursor 2nd is always the right one.

Cursors can not be positioned from the screen.

The displayed calculation results are continuously recalculated.

IMPORTANT NOTE:

In LOCK the TIME/DIV and AMPL/DIV settings can be changed, which is displayed in the top text area of the screen. These values are not valid for the locked picture on the screen.

The displayed cursor values are valid for the picture on the screen, so they don't correspond with the TIME/DIV and AMPL/DIV settings in the top text area.


The correct settings can be found in the bottom text area, if FULL text display is selected via the DISPLAY menu.

After performing other calculation functions, the results of these calculations will be displayed instead of the time or voltage values. If it is required to display the time or voltage values again, the selected calculation function has to be switched off by pushing the relevant softkey again.

Selected functions are always displayed with intensified text.

Registers which are not selected for display via the DISPLAY menu, are not shown in the CURSOR menu and can not be selected for cursor operation.

At power-down, the last selected cursor positions are saved in memory.



HA10140
10/15/02

Figure 4.19 Cursor display and control.

If a register has no trace, a message

Register has no valid settings: no cursors.

can be displayed.

- 1 R0 A B
- 2 R1 A B
- 3 R2 A B
- 4 R3 A B

If a register is selected for cursor operation, the cursors are visible on the screen for the selected (intensified) channel. The other channel can be selected by pushing the softkey once again. In the top text area the time (dt) and the voltage (dV) are displayed as described before.

The previous position of both cursors in a register is displayed when it is selected again.

In single channel operation, only the selected channel (for example R0 A) is visible. Other possible displays are (R.. ADD), or (R.. MUL).

If register R0 is selected for cursor operation, the text RESTART is visible next to softkey 7 and an automatic selection of time-base setting and/or delay is possible. (See explanation for softkey 7).

If A versus B was selected for the selected register, a message

Register in A versus B : no cursors possible.


is displayed and the cursors are switched off.

If a register is selected in which no traces and settings are stored, a message

Register has no legal settings : no cursors

is displayed.

5 CALCULATE>



If CALCULATE is selected, the CURSOR CALCULATE menu is displayed and various calculation routines can be selected.

The text CALCULATE is only visible if the cursors are switched on.

If ADD or MULTIPLY is selected, the dV value is given in dots (dts).

If EXT CLOCK is selected, the dt value is given in dots (d'ts).

5 1 RMS

After pushing RMS (Root Mean Square), the RMS value of the signal detail between the cursors is calculated. This RMS value is related to a zero level what is assumed to be at mid-memory.

The result is visible in the top text area by RMS: + V.



Figure 4.20 RMS measurement.

5 2 1/dt

After pushing 1/dt, the reciprocal time between the cursors is calculated. Only if the cursors are positioned on points, which are 1 signal period from each other, the displayed value is the frequency. The result is visible in the top text area by 1/dt: Hz. With both cursors placed over each other, 1/dt: 0/d is displayed.




Figure 4.21 1/dt measurement.


5 3 MEAN VALUE

After pushing MEAN VALUE, the mean value of the signal between the cursors is calculated. This mean value is related to a zero level what is assumed to be at mid-memory. The result is visible in the top text area by MEAN: V



Figure 4.22 Mean value measurement.

5 4 MULTIPLY>



If MULTIPLY is selected, the CURSOR MULTIPLY menu is displayed and the product of two traces within a selected register can be calculated and stored in another selected register.

In multiplication, mid-memory is used as 0 Volt reference. The maximum possible signal will be placed in the selected result register.

The execution of a multiply action takes place after pushing the softkey of the register where the result should be stored and goes on until it is stopped by deselecting the function.

5 4 1 R0 A*B
 5 4 2 R1 A*B
 5 4 3 R2 A*B
 5 4 4 R3 A*B

The register of which the traces have to be multiplied with each other can be selected with these softkeys.

The selected text is intensified.

If only one channel is present in the selected register, only R.. A*A or R.. B*B or R.. ADD*ADD or R.. MUL*MUL is displayed and the square of the signal can be calculated.

After selection of a register, the texts RESULT IN R1...R3 becomes visible.

5 4 5 R1
 5 4 6 R2
 5 4 7 R3

After selection of a register, one of the softkeys R1, R2 or R3 must be pushed in order to select in which register the result must be stored and in order to start the execution of the multiply action.

5 4 8 RETURN

After pushing softkey RETURN, the CURSOR CALCULATE menu becomes visible.

Softkey RETURN can also be used to avoid a multiply action, before a destination register is selected.

5 5 PEAK-PEAK

After pushing softkey PEAK-PEAK, the voltage between the lowest and the highest point of the signal between the two cursors, will be calculated and displayed in the top text area as PEAK: ... V.



Figure 4.23 Peak-peak measurement.

5 5 RISE TIME

After pushing softkey RISE TIME, the rise- or fall-time will be calculated and displayed in the top text area.

This means that the time between the first point on or above the first 10 % of the voltage between the cursors and the first point on or just above the first 90 % of the same voltage level or visa versa (as seen from the first cursor) is calculated.

The RISE time will be displayed in the top text area as RISE: + ... s and the FALL time as RISE:- ... s.



Figure 4.24 Rise time measurement.

5 7 LOCATE

The text LOCATE is visible, when PEAK-PEAK or RISE TIME measurement is selected.

After pushing softkey LOCATE, two markers are visible on the measured points:

At peak-peak measurement, one marker to the highest and one marker to the lowest value of the signal part between the cursors.

At rise-time measurement, one marker to the 10 % level and the other marker to the 90 % level of the signal part between the cursors.

Both sets of two markers can be visible at the same time.

The markers are suppressed again by pressing softkey LOCATE again or by switching of the PEAK-PEAK and the RISE TIME functions.

5 8 RETURN

After pushing softkey RETURN, the CURSORS menu will be visible again. The result of the last calculations is still displayed in the top text area.

6 CURSOR OFF

If the cursors are visible on the screen, an intensified text CURSOR OFF is visible. Pushing softkey CURSOR OFF switches the cursors off.

If the cursors are switched off, the calculation results in the top text area the markers on the trace and the text CURSOR OFF in the softkey area will disappear.

7 RESTART

If the cursors are in one of the traces of register R1, R2, or R3, the text RESTART is not visible. Pushing this softkey has no effect then.

If the cursors are in one of the traces of register R0 the text RESTART is displayed.

After pushing softkey RESTART an automatic horizontal acquisition expand is performed for the signal part between the two cursors.

A new trigger-delay setting (related to the first cursor) in DIV and a new and faster time-base setting (related to the time between the cursors) is automatically selected. After this selection a new signal is recorded with this new settings on the first active trigger.

The new settings are displayed in the top text area and the text RESTART is visible in high intensity.

The new settings are chosen in such a way, that the original signal part between the cursors, remains completely visible on the screen and is recorded with the maximum possible resolution.

If further expansion is not possible, this is indicated by one of the messages

**No further expansion possible!
RESTART not executed.**

or

**RESTART not fully executed because of
overflow in time base or trigger delay.**

NOTE: *The first message also appears, if the distance between the cursors is so big, that even in one time base setting faster, the original signal part between the cursors can't be displayed completely.*

8 REVERSE

Normally the text REVERSE is not visible.

After the first operation of the RESTART function, the text REVERSE is visible.

If softkey REVERSE is pushed, than the settings as before performing the RESTART action are recalled.



If pushbutton CLEAR is pushed and dots are selected for display, a zero line is written into the accumulator (R0) and no traces are displayed on the screen.

If DOTS is off or SMOOTH is selected for display and pushbutton CLEAR is pushed, a zero line is written into the accumulator (R0) and displayed on the C.R.T. screen.

The cleared contents of the accumulator can be saved in one of the other registers R1...R3, to clear the contents of these registers.

When pushbutton CLEAR is pushed during a SINGLE-shot, ■ MULTIPLE-shot or a ROLL-mode operation, the CLEAR pushbutton has a RESET function.

The function CLEAR is not possible in the LOCK mode.



If pushbutton SAVE/PLOT is pushed, the SAVE/PLOT menu is displayed. See 4.2.9.1.

4.2.9.1 SAVE/PLOT MENU STRUCTURE





Figure 4.26 Save/plot menu structure.

SAVE/PLOT MENU



After selection of the SAVE/PLOT menu by depressing pushbutton SAVE/PLOT, various SAVE and PLOT selections can be done. The register Rn which is selected for PLOT actions via the PLOT SELECT menu, is displayed on the screen.

- 1 --
- 2 SAVE TO R1
- 3 SAVE TO R2
- 4 SAVE TO R3

If a SAVE function is selected, register R0 is copied in the selected register. The SAVE action is performed as long as the softkey is pressed, so that register R0 is copied again after each acquisition.


A message

No direct SAVE in LOCK mode.

Press SAVE again during the display of this message.

becomes visible in the bottom text area for a short moment when the Instrument is in the LOCK-mode. A SAVE action is then only possible as long as the message is displayed.

5 ANALOG>



If ANALOG is selected, the contents of the displayed register Rn is transferred to the analog plot output. The default register is R0. Other registers to plot as well as the plot speed can be selected via the SELECT function of the SAVE/PLOT menu (softkey 8). During the PLOT action, the PLOT ANALOG menu is displayed and a dot, which moves from the left to the right (over 10 divisions) is displayed in the bottom text area of the screen to show the progress of the plot action and a message

***** PLOTTER ACTIVE *****

Changes are possible after plotter has stopped

is displayed.

At the end of the PLOT action, the menu SAVE/PLOT is displayed again.

- 5 1 --
- 5 2 --
- 5 3 --
- 5 4 --
- 5 5 --
- 5 6 --
- 5 7 --
- 5 8 STOP

A PLOT action can be interrupted by pushing softkey STOP. The SAVE/PLOT menu is then displayed again. During a plot action, it is not possible to switch to another softkey menu.

6 --

7 AUTO PLOT>

PLOT AUTOMATIC		1
—		2
—		3
—		4
ANALOG >		5
—		6
—		7
RETURN		8

AUTO PLOT functions only in the SINGLE-shot mode.

If AUTO PLOT is selected, the contents of register R0 is automatically transferred to an analog X-Y recorder.

The text AUTO PLOT is not visible in an other horizontal mode.

The contents of R0 is automatically plotted after each refreshment of the memory when a valid trigger is received.


7 1 —

7 2 —

7 3 —

7 4 —

7 5 ANALOG>



After selecting ANALOG, the AUTO.PLOT ANALOG menu is displayed and the function is active for an analog X-Y recorder.

The PLOT speed can be selected via the SELECT function of the SAVE/PLOT menu (softkey 8). During the PLOT action, the AUTO.PLOT ANALOG menu is displayed and a dot, which moves from the left to the right is displayed in the bottom text area of the screen and a message

***** PLOTTER ACTIVE *****
Changes are possible after plotter has stopped

is displayed.

At the end of the PLOT action, menu PLOT AUTOMATIC is displayed again.

```

7 5 1  --
7 5 2  --
7 5 3  --
7 5 4  --
7 5 5  --
7 5 6  --
7 5 7  --
7 5 ■ STOP

```

A PLOT action can be interrupted by pushing softkey STOP. The PLOT AUTOMATIC menu is then displayed again.

During a plot action it is not possible to switch to an other softkey menu.


```

7 6  --
7 7  --
7 8  RETURN

```

After pushing softkey RETURN, the SAVE/PLOT menu is displayed again. The selections as made before remain unchanged.

8 SELECT>



If SELECT is selected, a selection can be made for a register, a plot output speed and the penlift output polarity.

Only the registers which are selected for display, can be selected for plot actions.

8 1 R0


8 2 R1

8 3 R2

8 4 R3

Hard copies are made from the selected register.

8 5 ANALOG>



If ANALOG is selected, the PLOT SELECT ANALOG menu is displayed and the plot output speed as well as the penlift polarity can be selected.

8 5 1 PEN UP "0"

If the text PEN UP "0" is intensified, the instrument will generate a low level TTL signal on its PENLIFT output.

8 5 2 PEN UP "1"

If the text PEN UP "1" is intensified, the instrument will generate a high level TTL signal on its PENLIFT output.

8 5 3 —

8 5 4 —

8 5 5 UP

The actual analog plot output speed is displayed on the screen. This speed is always a multiple of 20 ms/dot. Pushing softkey UP increases the value in steps of

20 ms, which means that the actual plot output speed decreases. Above 200 ms/dot the value increases in steps of 100 ms/dot. The fastest possible output speed is 20 ms/dot.

If 2000 ms/DT is displayed, and UP is depressed, a message

PLOT TIME out of range. Range is 20 .. 2000 ms.

is displayed.

The real plot time may be different than selected, due to automatic plot control speed adaption related to the signal shape.

Note: This plot speed must be adapted to the speed of a relative slow X-Y recorder.

8 5 6 DOWN

The actual analog plot output speed is displayed on the screen. This speed is always a multiple of 20 ms/dot.

Pushing softkey DOWN decreases the value in steps of 20 ms, which means that the actual plot output speed increases. Above 200 ms/dot the value increases in steps of 100 ms/dot.

The slowest possible plot output speed is 2000 ms/dot.

If 20 ms/DT is displayed and DOWN is depressed, a message

PLOT TIME out of range. Range is 20 .. 2000 ms.

is displayed.

The real plot time may be different than selected, due to automatic plot control speed adaption related to the signal shape.

8 5 7 —

8 5 8 RETURN

After pushing softkey RETURN, the PLOT SELECT menu is displayed again. The selections as made before remain unchanged.

8 6 —

8 7 —

8 8 RETURN

After pushing softkey RETURN, the SAVE/PLOT menu is displayed again.

4.2.10. DISPLAY SECTION AND MENU STRUCTURE




Figure 4.27 Front panel view.



Continuously variable control giving vertical shift of one of the four registers R0, R1, R2, R3 (including channels in the registers) can be addressed via the DISPLAY menu.

The adjusting speed increases after turning continuously in one direction. After stopping, starting in the reverse direction resumes with a slow adjusting speed.



Continuously variable control giving horizontal shift of the total trace display.

Can also be addressed to one of the four registers R0, R1, R2, R3 (including channels in the registers) as long as menu DISPLAY Rn POSITION is selected.

The adjusting speed increases after turning continuously in one direction. After stopping, starting in the reverse direction resumes with a slow adjusting speed.



Continuously variable control for horizontal expand of the total trace display.

Can also be addressed to one of the four registers R0, R1, R2, R3 (including channels in the registers) as long as menu DISPLAY Rn POSITION is selected.

The adjusting speed increases after turning continuously in one direction. After stopping, starting in the reverse direction resumes with a slow adjusting speed.



Pilot lamp indicating that the X-EXPAND function is active for one or more registers. If the X-EXPAND control is fully turned anti clockwise, the UNCAL pilot lamp is switched off.



Selection between a display of discrete dots or a display of joined dots for all registers. DOTS can also be selected in A VERSUS B mode.



Selection between normal and smoothed display by switching a smoothing RC-filter in the display channel for all registers. The dots will disappear. SMOOTH can also be selected in A VERSUS B mode.

Note: SMOOTH can effect the signal on the screen, because of the decrease in bandwidth of the display channel.



If pushbutton MAGNIFY is pushed, the MAGNIFY menu is displayed. See 4.2.10.1.



If pushbutton DISPLAY is pushed, the DISPLAY menu is displayed. See 4.2.10.2.

4.2.10.1 MAGNIFY MENU STRUCTURE





Figure 4.28 Magnify menu structure.

MAGNIFY MENU



After selection of the MAGNIFY menu by depressing pushbutton MAGNIFY, the signal(s) can be expanded in horizontal or vertical direction.

*1, *2, *4, *8, *16 *32 or *64 horizontal expand can be selected.

Each part of the complete register can be displayed using the HORIZONTAL POSITION control.

1 EXPAND

With softkey EXPAND, higher horizontal expand factors can be selected until *64 is reached. Then the text expand is not displayed any longer. EXPAND is always with respect to mid-screen.

The selected factor is displayed in the second line of the softkey menu header.

With A VS B selected, the maximum expand factor is *8.

2 *1

With softkey *1, the horizontal expand factor is set to *1 which means –no expand–.

3 COMPRESS

With softkey COMPRESS, lower horizontal expand factors can be selected until *1 is reached. The selected factor is displayed in the second line of the softkey menu header. The text COMPRESS is not visible if the expand factor is *1.

4 —

5 Y*5

After pushing softkey Y*5, the vertical expand factor is set to *5 and 1/5 of the contents of the selected registers is displayed over 10 vertical divisions of which 8 divisions are visible. Mid-memory is displayed on the center-line of the graticule.

6 Y*1

Mit Softkey Y*1 wird die vertikale Dehnung auf *1 zurückgesetzt; der Inhalt der gewählten Register wird dann vertikal auf 10 cm angezeigt, von denen 8 cm sichtbar sind. Die Mitte des Speichers liegt auf der Mittellinie des Rasters.

7 Y/5

Nach dem Drücken von Softkey Y/5 wird der Inhalt von jedem Speicher vertikal auf 2 cm dargestellt. Wenn Y-POSITION sich in der Nullstellung befindet, sind die Register gleichmäßig über den Schirm verteilt.




Abb. 4.29 Schirmeinteilung bei Y/5.

4.2.10.2 DISPLAY, MENÜ-STRUKTUR

DISPLAY




Abb. 4.30 Display, Menü-Struktur.

6 Y*1

After pushing softkey Y*1, the vertical expand factor is set to *1 and the contents of the selected registers is displayed over 10 vertical divisions of which 8 divisions are visible. Mid-memory is displayed on the center-line of the graticule.

7 Y/5

After pushing softkey Y/5, the contents of each memory is displayed over 2 vertical divisions. If the Y-positions are in zero position, the registers are equally divided over the screen.




Figure 4.29 Y/5 display.

4-63E
REV. 1-77

4.2.10.2 DISPLAY MENU STRUCTURE






Figure 4.30 Display menu structure.

DISPLAY MENU



After selection of the DISPLAY menu by depressing pushbutton DISPLAY, various DISPLAY selections can be made.

1 REG SELECT>



After selecting REG SELECT, the DISPLAY/R SELECT menu is displayed and a selection can be made of the registers which have to be displayed.

1	1	R0	ON	OFF
1	2	R1	ON	OFF
1	3	R2	ON	OFF

1 4 R3 ON OFF

One or more registers can be selected for display.

Depending on the selection, the text ON or OFF is intensified. Pushing the relevant softkey changes the situation from ON to OFF or reverse.

If a register is switched OFF, the POSITION controls which are eventually addressed to this register, are from that moment on addressed to the next displayed register.

If a register which is not yet displayed, is selected for display, the POSITION controls are addressed to this register.

If cursors are selected, the cursors are on the selected register now.

1 5 --


1 6 --

1 7 --

1 8 RETURN

After pushing softkey RETURN, the DISPLAY menu is displayed again. The selected registers remain displayed.

2 POS Rn>



As long as softkey POS Rn is not pushed, the text POS Rn is displayed and the position of the selected register can be set with the VERTICAL POSITION and HORIZONTAL POSITION control. The HORIZONTAL POSITION control controls all other registers as well at the same moment. The X-EXPAND control functions also for the selected register.

After selecting POS Rn, the DISPLAY Rn POSITION menu is displayed and the Y-POSITION control, the X-POSITION control as well as the X-EXPAND control can be assigned to one of the four registers. The selected register number is intensified.

If a register is not selected for display, this register can not be selected for positioning.

The text Rn for this register is then not displayed.

2 1 R0

2 2 R1

2 3 R2

2 4 R3

The Y-POSITION, X-POSITION and X-EXPAND controls are addressed to the selected register.

2 5 CALIBRATE


After pushing softkey CALIBRATE, the traces are set as before pushing softkey CHANNEL (see next description).

The Y-POSITION is set so, that mid-memory is mid-screen in Y*1 and Y*5 mode and on +3, +1, -1 and/or -3 divisions in the Y/5 mode.

The X-POSITION is set so, that the beginning of the trace is situated on the most left graticule line.

The X-EXPAND is set to its calibrated position.

2 6 CHANNEL



If CHANNEL is selected, the POSITION Rn CHANNEL menu is displayed.

Traces within the same register can now be positioned in relation to each other.

If only one channel is recorded in the selected register, this menu can not be reached and the text CHANNEL is not visible.

2 6 1 CHANNEL A

2 6 2 CHANNEL B

The position of the selected channel A or B in the selected register Rn can be changed with the Y-POSITION, the X-POSITION and the X-EXPAND controls.

2 6 3 —

2 6 4 —

2 6 5 CALIBRATE

After pushing softkey CALIBRATE, the display modifications made under channel control are ignored and the traces are adjusted as follows.

The Y-POSITION is set so, that mid-memory is mid-screen in Y*1 and Y*5 mode and on +3, +1, -1, and/or -3 divisions in the Y/5 mode.

The X-POSITION is set so, that the beginning of the trace is situated on the most left graticule line.

The X-EXPAND is set to its calibrated position.

The controls are not longer addressed to one of the channels A or B but to the selected register.

2 6 6 —


2 6 7 —

2 6 8 RETURN

After pushing softkey RETURN, the DISPLAY Rn POSITION menu is displayed again.

All settings done before remain unchanged.

2 7 PLOT>



After selecting PLOT, the DISPLAY PLOT menu is selected and a plot can be made from the trace area of the screen. In this way a copy of the trace area can be made including the selected trace position changes.


2 7 1 --

2 7 2 --

2 7 3 --

2 7 4 --

2 7 5 ANALOG>



After selecting ANALOG, a copy of the screen is made on an analog X-Y recorder.

During the PLOT action, the PLOT ANALOG menu is displayed and a dot which moves from the left to the right (over 10 divisions) is displayed in the bottom text area of the screen to show the progress of the plot action and a message

***** PLOTTER ACTIVE *****
Changes are possible after plotter has stopped.

is displayed.

The settings made with the Y-POSITION control, the X-POSITION control and the X-EXPAND control remain.

At the end of the plot action the menu DISPLAY PLOT is displayed again.

```

2 7 5 1  —
2 7 5 2  —
2 7 5 3  —
2 7 5 4  —
2 7 5 5  —
2 7 5 6  —
2 7 5 7  —
2 7 8 8  STOP

```

A PLOT action can be interrupted by pushing softkey STOP. The DISPLAY PLOT menu is then displayed again.

```

2 7 6  —
2 7 7  —
2 7 8  RETURN

```

After pushing softkey RETURN, the DISPLAY Rn POSITON menu is displayed again.

```

2 8 RETURN

```

After pushing softkey RETURN, the DISPLAY menu is displayed again and the selected register (e.g. R2) can still be positioned. This is indicated by the text POS R2.

All modifications made between traces in X- as well as Y-direction (with the exception of Y changes, which are made via the CHANNEL menu), are eliminated.

3 INVERT>

DISPLAY INVERT	
R0 ON OFF	1
R1 ON OFF	2
R2 ON OFF	3
R3 ON OFF	4
---	5
---	6
---	7
RETURN	8

If INVERT is selected, the DISPLAY/INVERT menu is displayed and a selection can be made of the registers of which the display has to be inverted.

If a register is not selected for display, the inversion is not possible and the register number is not displayed.


3	1	R0	ON	OFF
3	2	R1	ON	OFF
3	3	R2	ON	OFF
3	4	R3	ON	OFF

One or more registers can be selected for Inverted display.
Depending on the selection, the text ON or OFF is intensified.
Pushing the relevant softkey changes the situation from ON to OFF or reverse.

3	5	---
3	6	---
3	7	---
3	8	RETURN

After pushing softkey RETURN, the DISPLAY menu is displayed again. The selections as made before remain unchanged.

4 A versus B >



If A versus B is selected, the DISPLAY A versus B menu is displayed. One or more registers can be selected for A versus B display. If a register is not selected for display, or if only one channel is recorded in a register, the text Rn ON OFF is not displayed and A versus B display is then not possible for this register. If no register is displayed with two channels, A versus B is not possible and the text A versus B is not displayed and the function can not be selected.

4	1	R0	ON	OFF
4	2	R1	ON	OFF
4	3	R2	ON	OFF
4	4	R3	ON	OFF

One or more registers can be selected for A versus B display. Depending on the selection the text ON or OFF is intensified. Pushing the relevant softkey changes the situation from ON to OFF or reverse. If the CURSORS were selected for the selected register, a message

Register in A versus B : no cursors possible.

is displayed and the cursors are switched off.

If a horizontal expand greater than *8 is selected via the MAGNIFY menu, a message


Max. horizontal expand in A versus B is *8

is displayed and horizontal expand is set to *8.

4	5	—
4	6	—
4	7	—
4	8	RETURN

After pushing softkey RETURN, the DISPLAY menu is displayed again. The selections as made before remain unchanged.

5 REDUCED➤



If REDUCED is selected, the DISPLAY REDUCED menu is displayed. Only in Y/5 one or more registers can be selected for REDUCED text.

The text REDUCED is only visible if Y/5 is selected via the MAGNIFY menu.

If a register is not selected for display, the text Rn ON OFF is not displayed and the action is not possible.




Figure 4.31 Reduced text display in Y/5.

PA11108
6/1/97

5	1	R0	ON	OFF
5	2	R1	ON	OFF
5	3	R2	ON	OFF
5	4	R3	ON	OFF


One or more registers can be selected for REDUCED TEXT display. Reduced text means that the parameters of the signal(s) of the selected register are displayed in the trace area. See also 4.2.2.

Pushing the relevant softkey changes the situation from ON to OFF or reverse.

5 5 —
 5 6 —
 5 7 —
 5 ■ RETURN

After pushing softkey RETURN, the DISPLAY menu is displayed again. The selections as made before remain unchanged.

6 FULL>



If FULL is selected, the DISPLAY/FULL TEXT menu is displayed.

The most important information of the front panel settings is always visible in the top text area but selection of the display of full information of a certain register or the additional information of the front panel setting, is possible.

If a register is not selected for display, the text Rn is not displayed and the action is not possible for this register.

If a register which is selected for full text is turned off via the REGISTER SELECT menu, full text is turned off.




Figure 4.32 Full text display.

6 1 R0

6 2 R1

6 3 R2

6 4 R3

The parameters of the selected register can be selected for display in the bottom text area.

Front information in the bottom text area, selected by softkey FRONT is not longer displayed then.

Pushing the relevant softkey changes the situation from ON to OFF or reverse.

6 5 FRONT

With softkey FRONT, two extra lines with front panel information are visible in the bottom text area.

Register information in the bottom text area, selected by one of the softkeys R0, R1, R2 or R3 is not longer displayed then.

6 6 --

6 7 TEXT OFF

This text is only visible if full text is selected.

The bottom text area is blanked when softkey TEXT OFF is pushed.

6 8 RETURN

After pushing softkey RETURN, the DISPLAY menu is displayed again. The selections as made before remain unchanged.

7 CHAN IDENT

After pushing softkey CHAN IDENT, the channel identification (A and/or B) on the C.R.T. screen can be switched on or off.

Pushing this softkey changes the situation from ON to OFF or reverse, indicated by a high or a low intensity of the text CHAN IDENT.

8 --

4.2.11. FRONT NR. SECTION AND MENU STRUCTURE




Figure 4.33 Front panel view.

FRONT NR.



If pushbutton FRONT Nr. is pushed, the FRONT Nr. menu is displayed.
See 4.3.11.1.

4.2.11.1 FRONT Nr. MENU STRUCTURE

FRONT Nr.





Figure 4.34 Front nr. menu structure.

FRONT NUMBER



If pushbutton FRONT No. is depressed, menu FRONT NUMBER will be visible on the screen and the internal front setting memory can be controlled.


FRONT No 00.00 is the actual front.

The last used front number is visible in the softkey text area and can be made actual again.


In this front setting memory, a number of different fronts can be saved and later on recalled. The number of fronts which can be programmed is displayed when an ENTER menu is selected.

When memory back-up batteries are installed, all saved fronts remain in memory after switching off the instrument.

Each front number contains a main front number and a sub front sequence number.



Example: FRO. 00.01



NAT 2437

Figure 4.35 Main and sub fronts.

By the RECALL, ENTER, NEXT and PREVIOUS actions, the actual front setting is saved in a one level front setting back-up memory.

If required, the front setting which is saved in the back-up memory, can be made actual again by entering <ENTER> <9999> <EXECUTE>.

After EXECUTE an automatic RECALL action is performed.

1 --

2 --

3 RECALL

If softkey RECALL is pushed, the front number which is displayed in the softkey text area is made actual.

After selection of the FRONT Nr menu, the displayed front number can be made actual after pressing softkey RECALL.

If the selected front number is not programmed, a message

**Selected front doesn't exist.
Front numbers not changed**

is displayed.

4 ENTER>

FRONT NMBR ENTER	
Memory for	1
77 fronts	2
Enter like:	3
main.sub	4
?	5
--	6
CLEAR	7
--	8
EXECUTE	9

If softkey ENTER is pushed, the FRONT NMBR/ENTER menu is displayed and a front number between FRO. 01.00 and FRO. 99.99 can be entered using the numeric keypad. The selected frontnumber is visible in the softkey text area.

4 1 --

4 2 --

4 3 --

4 4 --

4 5 --

4 ■ CLEAR

If an error is made, the selected front number can be cleared by pressing softkey CLEAR.

4 7 --

4 8 EXECUTE

By pressing softkey EXECUTE the selected front number is automatically recalled a RETURN is performed resulting in the display of the FRONT NUMBER menu. The number is tested against the real setting memory and if the number does not exist a message or not may be selected, one of the following messages

Selected front doesn't exist.
Front numbers not changed.

or

Front 0 cannot be selected because it is
the actual front panel setting.

or

Front number cannot go beyond 99.

or

Selected front doesn't exist.
Number set at last available front in sequence.

or

BACK UP FRONT cannot be changed. Action not executed.
 can be displayed.

5 NEXT

If softkey NEXT is pushed, the next higher sub frontnumber of the selected sequence is displayed in the softkey text area and made actual.

If NEXT goes to a number after the end of the sequence, the first front in the sequence will be made actual and a message

End of Sequence detected.
First front of sequence is selected.

is displayed.

If a front does not exist, a message

Selected front doesn't exist.
Front numbers not changed.

is displayed.

6 PREVIOUS

If softkey PREVIOUS is pushed the next lower sub front number of the selected sequence is displayed in the softkey text area and made actual.

It means that the front setting which is saved under this number, will be made actual. If PREVIOUS goes to a number before the begin of the sequence, the last front in the sequence will be made actual and a message

Begin of sequence detected.

Last front of sequence is selected.

is displayed.

If a front does not exist, a message

Selected front doesn't exist.

Front numbers not changed.

is displayed.

7 --

8 PROGRAM>

FRONT PROGRAM	
INSERT	1
DELETE	2
SAVE	3
ENTER >	4
FRO. 1.00	
UP	5
DOWN	6
End Of Seq	7
RETURN	8

If softkey PROGRAM is pushed, menu FRONT PROGRAM will be visible on the screen and new front setting sequences can be programmed or existing sequences can be altered.

By all DELETE and SAVE actions, the previous front setting is saved in a one level front setting back-up memory.

If required, the front setting which is saved in the back-up memory, can be made actual again by <ENTER><9999><EXECUTE> in the previous menu.

8 1 INSERT

If softkey INSERT is pushed, the actual front setting will be inserted after the front number which is displayed in the softkey text area. The number of the inserted front is displayed.

When the new front setting is inserted, all higher sub front numbers will be increased by one until the last front of the sequence is reached.




Figure 4.36 INSERT-action.

Messages

**Selected front doesn't exist.
Front numbers not changed.**

or

**Not enough memory to SAVE or INSERT front.
DELETE something before continue.**

can be displayed.

8 2 DELETE

If a programmed front setting should be removed from a sequence, the number of this front setting has to be selected first.

After pressing softkey DELETE, the selected frontpanel setting is removed from the sequence and all higher sub front numbers will be decreased by one.




Figure 4.37 DELETE-action.

If DELETE is operated by mistake it can be corrected by <ENTER> <9999> <EXECUTE> in the previous menu followed by an INSERT action with the correct front number.

A message

**Selected front doesn't exist.
Front numbers not changed.**

can be displayed.

NOTE: If a main front number is deleted, the complete front including all the belonging sub fronts is also deleted!

8 3 SAVE

If SAVE is depressed, the actual frontpanel setting can be saved under the number which is visible in the softkey text area.

If SAVE is operated by mistake it can be corrected by <ENTER> <9999> <EXECUTE> in the previous menu.

This can only be done if the front where save is done, already exists.




Figure 4.38 SAVE-action.

A message

**Not enough memory to SAVE or INSERT front.
DELETE something before continue.**

can be displayed.

8 4 ENTER>

FRONT NMBR
ENTER

Memory for 1

77 fronts 2

Enter like: 3

main.sub 4

? 5

-- 6

CLEAR 7

-- 8

EXECUTE 9

If softkey ENTER is pushed, the FRONT NUMBER/ENTER menu is displayed and a front number between FRO. 00.00 and FRO. 99.99 can be entered using the numeric keypad. The selected front number is visible in the softkey text area.

8 4 1 --

8 4 2 --

8 4 3 --

8 4 4 --

8 4 5 --

8 4 6 CLEAR

IF an error is made, the selected front number can be cleared by pressing softkey CLEAR.

8 4 7 --

8 4 8 EXECUTE

By pressing softkey EXECUTE the selected front number is entered and a RETURN is performed resulting in the display of the FRONT PROGRAM menu.

Entering a main front number is always possible.

Entering a sub front number is restricted to the following conditions:

- The number can not exceed the total number of fronts in the sequence.
A message

Selected front doesn't exist.

Number set at last available front in sequence.

can be displayed then and the last front number in the sequence is then selected.

8 5 UP

By pressing softkey UP, the next higher sub front number of the selected sequence is displayed in the softkey text area. UP functions until the EOS (end of sequence) is reached.

The following messages

**Selected front doesn't exist.
Front numbers not changed**

or

There is no sequence to this main front.

or

Begin or end of the sequence reached.

can be displayed.

8 6 DOWN

By pressing softkey DOWN, the next lower sub front number of the selected sequence is displayed in the softkey text area. DOWN functions until the first front is the sequence is reached.

The following messages

**Selected front doesn't exist.
Front numbers not changed.**

or

There is no sequence to this main frame.


or

Begin or end of the sequence reached.

can be displayed.

8 7 End of Seq

If softkey End of Seq. is pushed, the selected front number will become the last front number in the sequence and all the sub fronts after the selected front are deleted.



MAT 2441

Figure 4.39 End of Sequence-action.

A message

**Selected front doesn't exist.
Front numbers not changed.**

can be displayed.

8 8 RETURN

If softkey RETURN is pushed, the FRONT NUMBER menu will be displayed again.

REMOTE


Pilot lamp indicating that the instrument is in its REMOTE-state and that an (optional) interface overrides all the manually selected front panel settings. The last with the exception of POWER ON, ILLUM, TRACE ROT, INTENS TRACE, FOCUS and INTENS TEXT and PROBE INDICATION.

Resetting the instrument back to its LOCAL-state can be provided from a controller or by switching the instrument OFF and ON.

This REMOTE facility only functions in instruments which are provided with an (optional) interface.

Settings and outputs can then be controlled by other instruments external to the oscilloscope.

For an optional interface refer to the separate programming manual of this optional interface for the installation instructions and for the programming protocol.

URQ


The operator is able to ask an installed option for service by pushing the URQ (user request) pushbutton.

The kind of service which is given then depends on the user's program.

4.2.13

REAR PANEL




Figure 4.41 Rear panel view.

- 1 - Fan
- 2 - Space for IEEE connector.
- 3 - Mains input socket (90 V ... 264 Vac, 45 Hz ... 440 Hz). For safety instructions, read section 3.2.2.
- 4 - Mains fuse holder (Fuse rating 2.5 A delayed action). For safety instructions, read section 3.2.2.
- 5 - Space for RS-232C connector.
- 6 - Space for optional BNC sockets.
- 7 - Space for optional connector.
- 8 - Type plate with 12-number code and type number.
- 9 - Analog plot output socket.
- 10 - Memory back - up battery compartment with removable cover. For installation instructions, read section 3.3.

4.3 DETAILED OPERATING INFORMATION

4.3.1 Introduction

Before switching on, ensure that the oscilloscope has been installed in accordance with the instructions given in Chapter 3 and that the various precautions outlined have been observed.

The following sections give more detailed information regarding the specific functions of the instrument. It also gives a suitable starting routine before any measurements are made. Before reading this chapter, it is recommended that Chapter 4.2. Explanation of controls and sockets is first read.

This detailed information is especially useful for those operators who are not familiar with this type of oscilloscope.

The following subjects are described:

Start up procedure	4.3.2
Use of internal registers	4.3.3
Use of probes	4.3.4
Input coupling AC, 0, DC	4.3.5
OFFSET and SHIFT	4.3.6
Added mode and common-mode measurements	4.3.7
Triggering	4.3.8
Trigger delay	4.3.9
Time-base modes	4.3.10
MIN / MAX-mode	4.3.11
AVERAGE-mode	4.3.12
Horizontal magnifier	4.3.13
Vertical magnifier	4.3.14
A versus B mode	4.3.15
Analog PLOT-mode	4.3.16

4.3.2

Start-up procedure

Switch the instrument on, check that the power-on lamp is on and that the power-up routine is executed.

At the end of the power-up routine, the instrument is ready to accept input signals on the channel A and B input sockets.

Pushing pushbutton AUTO-SET will set the instrument for a triggered display on the C.R.T. screen. The INTENS controls can be set for a suitable trace and text intensity.

The channel A and B signals are stored now in register R0, which is one of the four internal registers R0, R1, R2 and R3, and the contents of register R0 are displayed. Eight of the ten vertical divisions are visible on the C.R.T. screen.




Figure 4.42 Screen of R0 contents in dual channel mode.

MSB 2011
43-903

Both signals can vertically be shifted over the screen using the (softkey selectable) OFFSET via the VERTICAL COUPLING menu or the input SHIFT (frontpanel control) facilities.

The signal amplitude and the number of displayed periods can directly be influenced by the AMPL/DIV controls and the TIME/DIV control. The selected settings can be read in the top text area on the C.R.T.

4.3.3 Use of internal registers

The procedure to save information in the internal registers or to clear the contents of these registers is now described.

Four internal registers R0, R1, R2 and R3 are available and in each of these registers a signal from channel A as well as a signal from channel B can be stored.

New signal information will always be stored in register R0 and the contents of this register can later be saved, if required, in one or more of the other three registers R1, R2 and/or R3 using the SAVE/PLOT menu by pressing the SAVE/PLOT pushbutton.

Using the DISPLAY menu by pressing pushbutton DISPLAY, gives the facility to the user to select one or more of the four internal registers to display their contents on the CRT screen. Then displayed signals are overlapping each other on the screen.

To avoid this, the MAGNIFY menu can be selected by pressing pushbutton MAGNIFY to select the Y/5 magnify mode which results in the following display of four registers.




Figure 4.43 Display of four registers in Y/5 mode.

W2243
80901

The memory system can be locked by depressing pushbutton LOCK and from that moment on no new signal acquisition can take place.

By depressing pushbutton WRITE, the system can be enabled again for new signal acquisition.

Register R0 can be cleared (in WRITE-mode only) using the CLEAR pushbutton.

The other registers can be cleared then by saving the cleared contents of register R0 in the selected register. In other words, clearing a register R1, R2 or R3 can be done by saving the blank contents of register R0 in the selected register.

4.3.4 Use of probes

To obtain a suitable connection between the oscilloscope and the device under test a probe is often used.

1:1 passive probes should only be used for d.c. and low-frequency signals. The capacitive loading of these probes attenuates high frequencies or increases the rise-time of signals to be measured (depending on the source impedance).

10:1 and 100:1 passive probes have smaller capacitive loading and are therefore better suited to measurements at higher frequencies. Against this, it may be a disadvantage that they decrease the sensitivity respectively by a factor of 10 and of 100.

For measurements at very high frequencies the high impedance probes are less suitable, due to reflections in the probe cable. This can be avoided by using probes with an input impedance of 5 k Ω and an attenuation factor of 100x. These probes have a very low capacitive loading, but a high resistive charge (5 k Ω).

FET active probes are superior, especially when measurements are to be taken from high impedance test points at the upper frequency limit of the oscilloscope bandwidth. Probes of this type have a very low capacitive loading, yet do not reduce the sensitivity.

10:1 and 100:1 passive probes must be properly compensated before use. Incorrect compensation leads to pulse distortion or amplitude errors at high frequencies.

For correct adjustment of the 10:1 attenuator probes delivered with this oscilloscope, refer to Section 8.1.1.4.

For more details about the available PHILIPS probes contact your local PHILIPS sales organisation.

4.3.5 Input coupling AC, 0, DC

The signals under observation are fed to input socket A and/or B and the AC, 0, DC coupling is set to either AC or DC depending on the compensation of the signal. As the vertical amplifier is d.c.-coupled, the full bandwidth of the instrument is available, and d.c. components are displayed as trace shifts if DC is selected. This may be inconvenient when small signals superimposed on high d.c. voltages need to be displayed. Any attenuation of the signal will also result in attenuation of the small a.c. component. This is remedied by using AC coupling, which employs a blocking capacitor to suppress the d.c. and l.f. signals. However, some pulse droop will occur when i.f. square-wave signals are displayed.

The 0 coupling interrupts the input signal and grounds the amplifier input, thus providing a quick check of the 0 V reference level.

4.3.6 OFFSET and SHIFT

The OFFSET (before attenuator) facility, which can be operated via the VERTICAL COUPLING menu, can be used to bring channel A and channel B roughly at the same DC level, so that a better use can be made of the vertical resolution.

The SHIFT (after attenuator and before memory) frontpanel controls can be used to adjust the channels against each other with an even better resolution.

4.3.7 Add mode and common mode measurements

If the ADD mode is selected, the signal voltages of both vertical channels are added. Depending on the selection of the INVERT-mode of the A or B channel, either the sum or the difference of the input signal is displayed. The ADD mode also enables differential measurements to be made. With these measurements, advantage is taken on the common mode rejection in the ADD mode. When the channel B is selected for inversion (and channel A INVERT not activated), the common mode portions of the signals on input sockets A and B will only be subjected to very slight amplification compared with the differential mode portions (see Fig. 4.44). In measurements where signal lines carry substantial common mode signals, e.g. hum, the differential mode will largely cancel out these signals and leave the remainder of interest (A-B). The capability of the oscilloscope to suppress common mode signals is given by the Common Mode Rejection (CMRR) factor. To obtain the degree of common mode rejection as specified, channel A and B gains must first be equalised. This can be done by connecting both channels to the CAL output connector, and adjusting one of the vertical VARIABLE controls for minimum deflection on the screen.

When passive attenuator probes are used, a similar equalisation process is recommended by adjusting their compensating control for minimum deflection.




Figure 4.44 Suppression of common mode signals.

4.3.8 Triggering

Trigger sources

The acquisition system can be triggered on signals which are derived from the vertical channels A and B. If two time-related signals are displayed via channels A and B, a stable display is obtained if the trigger source is the signal with the lower frequency.

The acquisition system can also be triggered on a signal which is applied to the external trigger input socket EXT TRIG. This input can be set for two different sensitivities EXT or EXT:10.

If LINE triggering is selected, the trigger signal is directly derived from the mains voltage and has the same frequency as the mains voltage. This trigger mode is useful, for instance, when examining the mains voltage ripple on the d.c. output voltage of a power supply.

Trigger filters.

Trigger filters can be applied for the selected trigger source. The following filter modes can be selected: DC, AC, LF Reject, HF Reject and HF Reject + DC.




Figure 4.45 Trigger filter responses.

- With DC selected, the full bandwidth of the trigger channel is available.
- With AC selected, the lower frequency range of the trigger channel is limited to 10 Hz. This is achieved by introducing a d.c. blocking capacitor into the signal path. This mode is recommended when triggering on an a.c. signal that is superimposed on a high d.c. voltage.
- With LF Reject selected, the trigger signal path is a.c.-coupled and signals with a frequency lower than 50 kHz are blocked by a high-pass filter. This mode is useful for triggering on an h.f. signal that is polluted with l.f. noise (e.g. hum).
- With HF Reject selected, the trigger signal path is a.c.-coupled and signals with a frequency higher than 50 kHz are blocked by a low-pass filter. This mode is advisable by high-frequency noise.
- With DC and HF Reject both selected, the signal path is d.c.-coupled and provided with a low-pass filter which blocks signals with a frequency higher than 50 kHz.

LEVEL and NEGATIVE SLOPE controls

The position of the LEVEL control determines the starting point of the signal acquisition. The circuit behind the LEVEL control functions as follows:

The trigger signal is fed to the input of a comparator and the voltage on the other input of this comparator is determined by the position of the LEVEL control.

If the trigger signal reaches the voltage level of the LEVEL control, a trigger pulse is generated by the comparator and the signal acquisition starts.

In this way, signal acquisition is started at a fixed point of the trigger signal, which means that by using the LEVEL control, it is possible to scan the shape of the waveform.

The NEGATIVE SLOPE pushbutton permits selection of the trigger slope.

If NEGATIVE SLOPE is selected, the signal acquisition starts on the negative-going slope of the trigger signal, otherwise on the positive-going slope.

If the LEVEL control is turned higher or lower than the maximum resp. minimum level of the selected input signal, the internal comparator does not generate a trigger pulse. The led "NOT TRIG'D" on the frontpanel will then light. In AUTO level triggering, the trigger level can not be turned "out of range".

The LEVEL and NEGATIVE SLOPE controls enable the trigger level to be set at a predetermined value without the need of an input signal. This is important when the signal to be measured is not available in advance as when single phenomena are to be observed. The selected trigger level point is indicated on the left side of the C.R.T. screen by one of the marks \uparrow , \downarrow , or \times . When input signals which exceed a known trigger level have to be displayed, the trigger level can be set in advance and an input signal of sufficient amplitude will initiate the recording of the signal.

When external triggering is selected, the \downarrow is replaced by an X. For line triggering an L is used as trigger level indication. Both the X and the L do not indicate the exact trigger level, related to the signal, but give an indication where the trigger level is set.

AUTO-triggering:

The acquisition system becomes free running 100 ms after the last trigger pulse. This means that even during the absence of trigger pulses the R0/ contents is refreshed, although the pilot lamp NOT TRIG'D lights. If a trigger is again applied, normal triggering is obtained. The lamp NOT TRIG'D will extinguish.

An additional feature of the AUTO mode is that the LEVEL range lies within the peak-to-peak value of the signal amplitude.

Due to the fact that the system becomes free-running 100 ms after the last trigger pulse, the AUTO-mode can not be used for signals with a repetition rate of 100 ms or more; for this the triggered mode must be selected.

DUAL triggering

By selecting DUAL triggering, the oscilloscope is able to trigger on either positive or negative-going edges of the signal. This feature can very well be used in the SINGLE-shot mode.

4.3.9

Trigger delay

TIME DIV positive trigger delay:

Range in 360 s/div ... 200 ns/div : 1 ... 9999 div

Range in 100 ns/div ... 5 ns/div : 1 ... 500 div

With the trigger delay facility the time (in time or divisions) between triggering and the start of the C.R.T. display (left-hand side) can be selected.

Example:

Suppose the 6th line in a TV pattern is desired (TV line = 64 us).

The required trigger delay is therefore $5 \times 64 \text{ us} = 320 \text{ us}$ (i.e. after the 5th line has been passed).

- Select TVF triggering.
- Set the TIME/DIV switch to position 10 us/div.
- Select a trigger delay of 320 us or 32 divisions.

A delay of 320 us between the frame pulse and the left-hand side of the C.R.T. display is now obtained.

This results in a display of the information of the 6th line.

TIME DIV negative trigger delay:

Range in 360 s/div ... 5 ns/div : -10 div

With negative trigger delay (pretrigger), selected, the trigger point can be chosen all over the ten horizontal divisions of the CRT screen. This enables to look at what happened before the trigger moment.

NOTE: *If the TIME/DIV switch is set to another position, the setting of the trigger delay positive or negative and in time or divisions, will be automatically changed (recalculated) and displayed in the top text area. The result of this recalculation is rounded off downwards to whole divisions.*

EVENTS trigger delay

Range : 1 ... 9999 events

In this mode which can be selected via the TRIGGER DELAY menu, the trigger can be delayed by a number of events before a new signal acquisition is started.

These events have to be applied to the EVENTS/EXT CLOCK BNC input socket on the frontpanel. The selected level on which these events pass is indicated in the TRIGGER COUPLING EV/EXT CLOCK menu. It is also indicated by a mark E on the left side of the C.R.T. Two predetermined values can be selected by softkeys: TTL and ECL level.

The number of events to be counted can be selected via the TRIGGER DELAY menu. Event counting is started at the arrival of a trigger signal.




Figure 4.46 Triggering after a number of events.

W0001
W001

4.3.10 Time-base modes

4.3.10.1 RECURRENT mode (5 s/div...5 ns/div)

This is the normal time-base mode in which the contents of register R0 is continuously refreshed on the receipt of new incoming trigger signals.

The RECURRENT mode can be divided in two ranges:

- Direct mode (5 s/div...200 ns/div)

From the incoming signals samples are taken real time and converted into digital codes which are placed in register R0

- Random sampling mode (100 ns/div ... 5 ns/div)

Incoming signals are converted into digital codes by using a time converting sampling method.
Only repetitive signals can be converted, which is indicated by the pilot lamp REP ONLY.

Direct mode single channel

In the time base range 5 s/div ... 1 ms/div, a sweep consists of 4096 samples, which are placed in register R0.

The first 4000 samples are displayed over 10 divisions on the screen.

In the time base range 500 ns/div ... 200 ns/div a sweep consists of 512 samples, but register R0 still contains 4096 locations.

The samples are placed on register locations 0, 8, 16, 24, etc., The other locations are filled with "samples" which are calculated by means of interpolation.

The interpolated samples are not displayed, if DOTS is on. This can be made clearly visible by using the *64 HORIZONTAL EXPAND factor via the MAGNIFY menu.

Direct mode dual channel

In dual channel mode, register R0 consists of 2x 2048 locations. (2048 locations per channel). Now, 2048 samples are taken simultaneously on each channel and placed in register R0. The first 2000 samples of each channel are displayed over 10 divisions on the screen. In the time base range 500 us/div ... 200 ns/div a sweep consists of 2 x 512 samples. Now the samples are placed on register locations 0, 4, 8, 12 and so on. The other locations are filled with interpolated "samples" and not displayed if DOTS is on.

In the time base range 500 us/div ... 200 ns/div, the interpolated samples can be replaced by real samples if the input signal has a repetitive character. This is called MAXIMUM RESOLUTION. If MAXIMUM RESOLUTION is selected, the pilot lamp REP ONLY is on.

In single channel mode the first sample of a sweep can be placed on location zero, as well as on location seven. To determine on which location the first sample should be placed, the time which elapses between the trigger pulse and the first sample pulse is measured by the internal delta-t circuit. A very short time results in placing the first sample on the first location, a longer time results in placing the first sample in a higher location. (see fig 4.47).

During the next sweep, again 512 samples are taken and placed in the register. If the samples are placed on locations where already has been placed a sweep of samples, the old samples are overwritten.

If AVERAGE is selected, the old sample and the new sample are used to calculate a new sample (see also section 4.3.12).

So it takes at least 8 sweeps, but probably more, to make a complete picture.




Figure 4.47 Maximum resolution mode after two sweeps.

In dual channel mode there are four possibilities to place the sweeps in register R0. Once again the time that elapses between the trigger pulse and the first sample pulse is used to place the samples at the right locations.

If AVERAGE is selected in the time base range 500 us/div ... 200 ns/div, the instrument automatically selects the MAXIMUM RESOLUTION mode.

When MIN / MAX is selected, the MAXIMUM RESOLUTION mode is automatically switched off.

NOTE: *In maximum resolution mode at least 8 sweeps (or 4 in DUAL channel mode) are needed to make a complete picture. In this way, a high resolution display with low jitter is realised.*

This also means that, if the input signal changes, there is temporarily a situation of change over, which looks quite strange. This effect becomes more strongly visible when the AVERAGE mode is turned on. Turning SHIFT, VARIABLE or LEVEL, has a similar effect.

After a change of an AMPL/DIV or TIME/DIV setting the register is cleared and a new picture is made. So there are no strange effects visible.

Random sampling mode

In the time base range 100 ns/div...5 ns/div, the random sampling mode is used. This mode allows the digitizing of input signals with high frequencies.

These frequencies may be even higher than the sampling and conversion rate. It is possible to digitize a 200 MHz signal with the sample and conversion rate of 50 Msamples/s.

To get a complete picture one condition should be met: the signal has to be repetitive.

The principle is as follows (see fig. 4.48):

After a trigger occurs, the time which elapses between the trigger pulse and the first sample pulse is measured by the delta-t circuit. Afterwards, a number of samples, depending on the time base setting, is taken and converted.

E.g. with the sampling rate of 50 Ms/s every 20 nsec a sample is taken. With a time base setting of 10 nsec/div a sweep takes 100 nsec. So $100/20 = 5$ samples can be taken during one sweep.

The locations in the register, where these samples have to be stored depend on the measured time between the trigger pulse and the first sample pulse.

The next sweep another 5 samples are taken and placed in register R0 and on the C.R.T. screen. Because there is probably another time difference between the trigger pulse and the first sample pulse, the samples are placed on other places in the register and so they make the picture more complete. To make a complete picture of 500 dots, at least 100 sweeps (in this example) are needed, but probably more. The more dots that there are already filled in, the bigger the chance is that a new sweep overwrites a previous sweep.

A complete picture consists of 500 dots over 10 divisions (512 samples in register R0) in single channel mode and of 2x500 dots over 10 divisions (2x512 samples in register R0) in dual channel mode.


If DOTS is off intermediate dots are calculated and displayed in the same way as in the direct mode with the MAXIMUM RESOLUTION mode off.

Because there are samples taken and stored in the oscilloscope before the trigger pulse occurs, pretrigger is also possible.


The urgency of a repetitive signal as input signal will be clear now.

NOTE: *When the input signal changes, the picture on the screen changes in a random way from the old situation to the new one. This looks quite strange for a while, but it is normal because of the random sampling principle.*


Because there is no time relation between the trigger pulses, which are derived from the input signal, and the sampling pulses, the time between a trigger pulse and the first sample pulse is random. This is why this principle is called random sampling.




AFTER ONE SWEEP



AFTER TWO SWEEPS



AFTER THREE SWEEPS



AFTER SWEEPS

NIST 2448

Figure 4.48 Random sampling.

4.3.10.2 SINGLE and MULTIPLE-modes

SINGLE and MULTIPLE shots can be taken in the direct mode (time base range 5 s/div ... 200 ns/div). This is very useful for displaying single phenomena.

SINGLE and MULTIPLE scans can be taken in the random sampling mode (time base range 5 ns/div ... 100 ns/div), and in the direct mode when the MAXIMUM RESOLUTION mode is on (time base range 500 us/div ... 200 ns/div). Then the pilot lamp "REP ONLY" is on. During a scan there are taken as many sweeps as necessary to make a complete picture on the screen.

When the SINGLE-mode is selected, the contents of register R0 is refreshed once after a trigger pulse and the selected delay-time, and the refreshed contents of register R0 is displayed on the C.R.T. screen.

As long as the instrument is waiting for a trigger pulse, the pilot lamp NOT TRIG'D will light.

When the MULTIPLE-mode is selected, the previously described SINGLE-action is repeated four times. The result of the first SINGLE action in register R0 is copied to register R3, the second result is copied to register R2, the third to register R1 and the fourth stays in register R0. This is independent of the registers being displayed or not.

The number of single actions to be executed is counted down in the top text area on the C.R.T.

When either the SINGLE- or the MULTIPLE-action is not completed, the text ARMD (ARMeD) in the soft-key text area is intensified.

A new shot or scan can be taken by pressing the relevant softkey or the CLEAR pushbutton.

In the time base range 5 s/div...1 ms/div, the resolution is 1x4096 or 2x2048 samples (the same as in RECURRENT mode).

In the time base range 500 us/div...200 ns/div, the resolution is 512 samples per channel at each shot.

The samples are placed in the locations 0, 8, 16, 24, etc. of register R0. Intermediate samples until 1x4096 or 2x2048 are calculated and not displayed if DOTS is on.

In the time base range 100 ns/div...5 ns/div, a scan takes as many sweeps until 512 samples, or 2x512 samples, are taken. Intermediate samples until 1x4096 or 2x2048 are calculated and not displayed if DOTS is on.

4.3.10.3 ROLL mode

The continuous ROLL mode is typically used for very low frequency signals and is effective with TIME/DIV settings from 50 ms/div... 360 s/div. The signal is recorded in register R0 and is built-up point-by-point from the right-hand side of the C.R.T. screen and continuously shifted towards the left.

This mode can be selected via the HORIZONTAL MODE menu and can be started by pressing the RUN/STOP softkey once.




Figure 4.49 ROLL-mode action.

No trigger is needed in this mode.

During the ROLL mode action, the text RUN in the softkey text area is intensified displayed. The ROLL mode can be stopped by pressing the RUN/STOP softkey once again. The text STOP in the softkey area is then intensified displayed.

When TRIG STOP is selected, the ROLL mode can be stopped by a trigger signal. The triggered STOP can be delayed by selecting a trigger delay.

The resolution is 4000 dots over 10 divisions in single channel mode and 2000 dots (per channel) over 10 divisions in dual channel mode.

The ROLL mode can be started again by depressing the RUN/STOP softkey or pushbutton CLEAR.

NOTE: In AUTO trigger mode every 0,1 s a trigger pulse is generated which gives a triggered stop.

If trigger delay 0 is selected, the trigger (stop) point is at the left side of the screen. This means that after the actual stop signal the screen continues 10 divisions.

4.3.10.4 EXT CLOCK mode

By applying an external clock frequency of max. 50 kHz to the EVENTS/EXT CLOCK BNC input socket and with the pushbutton EXT CLOCK depressed, the user is able to determine the conversion rate.


The internal digital time-base generator is not used then, which is indicated by the text TB EXTERN in the top text area on the C.R.T. screen.

The trigger delay facility still functions but can only be selected in divisions.

Cursor time measurements are not possible, which is indicated in the top text area by the text EXT.CLOCK!! in stead of dV....

4.3.11 MIN / MAX mode

In this vertical processing mode which can be selected in the time-base positions 5 s/div...5 us/div and in the ROLL mode, the maximum and minimum signal amplitude over the time between two adjacent input signal samples is measured by two PEAK DETECTORS per channel. These maximum and minimum signal levels are applied in serial to a TRACK AND HOLD gate via a MULTIPLEXER. This MULTIPLEXER switches between the minimum and maximum PEAK DETECTOR output. This results in peak values MIN - MAX - MIN - MAX - MIN and so on. The output values of the TRACK AND HOLD gate are applied to the ANALOG TO DIGITAL CONVERTER, converted, stored in the digital trace memory and displayed on the C.R.T. screen.




MAT 2451

Figure 4.50 Principle of peak detection.

After the moments that a minimum or a maximum is taken over by the TRACK AND HOLD gate before the ADC, the relevant PEAK DETECTOR will be reset.

During this reset cycle (about 20 ns) the PEAK DETECTOR is not able to watch the input signal.



MAT 2452

Figure 4.51 Principle of peak detection.

In the time base setting 5 usec/div the sweplelength is 50 us.

A sweep consists of 250 MIN and 250 MAX samples over 10 divisions. So the time between 2 MAX samples is $50/250 = 0,2 \text{ us} = 200 \text{ ns}$. In this 200 ns the PEAK DETECTOR is during 20 ns blind, due to reset. The blind time ratio is now 10%.

In faster timebase settings, the blind time ratio increases; MIN / MAX can not be used then.

In time base setting 50 us/div, the blind time ratio is 1%.

In time base settings 500 us/div and slower, the blind time ratio is 1°/oo or better.

The MIN / MAX mode can be selected for:

- Glitch detection
- Envelope measurements
- Detection of the aliasing

Glitch detection

Glitch detection in a digital storage oscilloscope is usually not possible. By means of the MIN/MAX mode, even glitches with a pulse width of 3 ns are still displayed with an amplitude of about 50%.

If a glitch occurs in the reset time of a PEAK DETECTOR it is not seen. Glitches longer than 20 nsec are always seen. The amplitudes depend on the part of the glitch that is outside the reset time of the PEAK DETECTOR. The chance of not catching a glitch can be decreased by the selection of a lower time base frequency.

Envelope measurements

Because the minimum and maximum value of the input signal is measured between every two adjacent samples, the MIN/MAX mode is perfectly suitable to measure and to display the envelope of an Amplitude Modulated r.f. signal.

Detection of the aliasing

In a digital storage oscilloscope the input signal is sampled at a high frequency, to convert the analog signals into digital signals. This sampling frequency is determined by the time-base setting or by the EXT CLOCK input frequency.

If the frequency of the input signal is almost the same (or a multiple higher) as the frequency of the sampling clock, a low frequency interference signal will be displayed, which seems to be not triggered on the screen, but the NOT TRIG'D pilot lamp does not light.

This phenomena is called aliasing.

Usually this is indicated by the ALIASED pilot lamp.

This lamp lights if the sampling frequency is smaller than twice the trigger pulse frequency. The latter is usually twice the selected input frequency.

Sometimes aliasing might still happen, but is not indicated. (E.g. an AM-signal, while the oscilloscope is EXT triggered on the LF-component).

To determine if the displayed signal is correct, the MIN/MAX mode can be selected. If the envelope of the input signal is now displayed, there was aliasing.

The chance of aliasing can be reduced by selecting an as fast as possible time base setting.

4.3.12 AVERAGE mode

AVERAGING is a way to suppress noise without losing bandwidth and can only be used in RECURRENT mode.

Every dot is calculated after every sweep in the following way:

$$\text{new} = \text{previous} + \frac{\text{measured} - \text{previous}}{C}$$

In this formula "previous" is a sample on the same position of the previous sweep.

If C=1 every new dot is the measured dot; AVERAGE is OFF. If C>1, the dot positions change slower.

The bigger C is, the slower the dot positions change. The following values for C can be selected: C = 2, 4, 8, 16, 32 or 64.

In ROLL mode it is also possible to use AVERAGE.

Between two displayed samples, more samples are taken. These intermediate samples are used to perform the AVERAGE calculation with a fixed value for C of C = 32.

AVERAGE cannot be used together with MIN / MAX.

NOTE: If SINGLE or MULTIPLE SHOT is selected AVERAGE is not effective.

4.3.13 Horizontal magnifier

When the continuous horizontal expand (X-EXPAND) control is used, the display on the screen expands horizontally to more than 2x the TIME/DIV setting. The reduced time window provides a more detailed display. Using the X POSITION control, any portion of the trace can be shifted into the display area.

Via the MAGNIFY menu the following horizontal expand factors can be selected: *1, *2, *4, *8, *16, *32 and *64.

4.3.14 Vertical magnifier

Via the MAGNIFY menu, three different vertical expand factors, Y/5, Y*1 and Y*5 can be selected.

Y/5

All four registers R0, R1, R2 and R3 can be displayed in their own trace area, each with a vertical trace height of two divisions.




Figure 4.52 Y/5 mode.

In this mode it is also possible to add reduced register texts in the trace area of each register via the DISPLAY menu.

Y*1

This is the normal vertical expand factor, in which each register can be displayed over 10 vertical divisions of which 8 divisions are visible on the C.R.T. screen.




Figure 4.53 Y*1 mode.

4-109E
8-10-77

Y*5

In this mode a vertical expand from 10 divisions to 50 divisions is possible of which only 8 divisions are visible on the C.R.T. screen.

Using the Y position control, any vertical portion of the trace can be shifted into the trace area of the C.R.T. screen.




Figure 4.54 Y*5 mode.

4-109E
8-10-77

4.3.15 A versus B-mode

With a A versus B selected, XY display is obtained from the samples derived from the channel A and channel B inputs.

The samples derived from the channel A input signal are used for horizontal deflection and the samples derived from the channel B input for vertical deflection.

This A versus B mode is principally a different mode than the real time X-Y mode in a real time oscilloscope.

In this case it is only a different way of displaying the contents of the registers. The storage of signal information is influenced by the position of the TIME/DIV switch, trigger selection and trigger SLOPE and LEVEL, so the A versus B display of this information is also influenced by these factors.

4.3.16 Analog plot mode

Two different PLOT-modes can be selected:

- Register PLOT via the SAVE/PLOT menu:

The contents of the selected register can be plotted. The Y-POSITION, the X-POSITION, and the X-EXPAND controls and the MAGNIFY selection have no influence.

- Screen PLOT via the DISPLAY menu:

The picture on the trace area of the C.R.T. screen can be plotted including the influence of the Y-POSITION, the X-POSITION and the X-EXPAND controls and the MAGNIFY selections. The picture will be similar to the picture on the screen.

The X and Y plot outputs on the rear panel of the oscilloscope generate 1 V full memory or 1 V full screen. (See characteristics).

The PEN LIFT output is an open collector output; max 12 V (TTL compatible).

Via the SAVE/PLOT menu the following selections can be done.

- PEN UP "0"
- PEN UP "1"
- PLOT speed
Range: 20 ms/dot ... 2000 ms/dot.
- Register R0, R1, R2 or R3 (only for register PLOT).


During plotting, the oscilloscope is in the LOCK-mode, which means that the contents of all registers cannot be changed.

In case of dual channel plotting, first channel A will be plotted and then channel B.

The PLOT operation is provided with a short delay at the start and end of the action to give sufficient time for manual pen positioning if no automatic pen lift function is available on the recorder.

If the SINGLE-shot mode and register plot is selected, the AUTO PLOT function can be activated. The contents of register R0 is then automatically plotted after each refreshment of the register when a valid trigger is received.

The progress of the plotting is indicated by a dot on the C.R.T. screen, which moves from the left to the right and which is displayed in the bottom text area.



4.4 PRINCIPLE OF OPERATION

In this section, the principles of operation are discussed at block diagram level, with special emphasis being applied to those parts of the circuit that differ from normal oscilloscope practice, i.e. the digital storage and control facilities.

4.4.1 GENERAL

This digital storage oscilloscope comprises the following sections:

- a signal acquisition section which can be divided in a vertical section, a trigger section and a conversion and processing section.
- a storage section
- a display section
- a control section
- a front section
- a power supply section

4.4.2 The signal acquisition section

Vertical section

The input signal(s) to be displayed can be applied to the BNC input sockets A (and) or B.

If the probe delivered with the oscilloscope is used, the attenuator setting of the oscilloscope is adapted to the attenuator factor of the probe via the PROBE DETECTION contacts adjacent to the BNC input connectors A and B.

The channels A and B are identical so only channel A is described.

The applied input signal is fed to the VERTICAL SIGNAL CONTROL via ATTENUATOR A. In ATTENUATOR A, the vertical sensitivity is determined, controlled by the AMPL/DIV button via the CONTROL SECTION (indicated with C →).

With the pushbutton COUPLING, a softkey menu is activated which makes selection of the following functions possible:

- input coupling selection : AC-DC-GROUND
- input impedance selection: 1 MOhm - 50 Ohm
- OFFSET selection (via preselected values)

The attenuation can be continuously controlled by the frontpanel VARIABLE control and when active the UNCAL led will be on.


Vertical shift of the trace is determined in the VERTICAL SIGNAL CONTROL by the frontpanel SHIFT control.

The pushbutton MODE makes via the softkeys selection of the following functions possible:

- INVERT A (B) : inversion of input signal
- VERTICAL DISPLAY MODES :
 - A : only channel A
 - A and B : channels A and B
 - ADD : A and B added
 - B : only channel B
- BW LIMIT : bandwidth limiter
- MIN / MAX : minimum / maximum mode

The VERTICAL DISPLAY MODE is selected in the VERTICAL CHANNEL SWITCH which also contains peak detectors to perform the softkey selectable MIN / MAX function.

The selected vertical input signal is then fed to the CONVERSION AND PROCESSING SECTION.



4.4 PRINCIPLES

In this section, the emphasis is on the principles of the system.

4.4.1 GENERAL

This diagram illustrates the general principles of the system.

- a sign.
- convey.
- a stor.
- a disp.
- a con.
- a fron.
- a pow.

4.4.2 The sign

Vertical

The input

If the program is adapted to the BNC input, the character of the signal in the ATTEN is the CON.

With the following:

- Input
- Input
- OFFS

The active th Vertical: SHIFT c

The push

- INVEF
- VERT

- BW L
- MIN /

The VEF contains Thesele




Figure 4.55 Blockdiagram.





Abb. 4.55 Blockschaltbild.

E SECTION



4.4 FUNKTIONSPRINZIP

In diesem Abschnitt wird die Arbeitsweise des Geräts im Blockschaltbild beschrieben, wobei vor allem auf diejenigen Teile der Schaltung eingegangen werden soll, die von einem normalen Oszilloskop abweichen; die digitale Speicherung und die Bearbeitungsmöglichkeiten.

4.4.1 ALLGEMEINES

Dieses digitale Speicher-Oszilloskop enthält folgende Baugruppen:

- einen Signalerfassungs-Teil, der in einen vertikalen Teil, einen Trigger-Teil und einen Teil für Umsetzung und Verarbeitung unterteilt werden kann;
- einen Speicherteil;
- einen Anzeigeteil;
- einen Steuerteil;
- einen Front-Teil;
- eine Stromversorgung.

4.4.2 Der Signalerfassungs-Teil

Vertikaltell

Die darzustellenden Signale können an die BNC-Eingangsbuchsen A und/oder B angeschlossen werden.

Wird der mit dem Oszilloskop mitgelieferte Tastkopf verwendet, dann wird die Abschwächer-einstellung des Oszilloskops mit Hilfe der PROBE DETECTION-Kontakte neben den BNC-Eingangsbuchsen A und B an den Abschwächungsfaktor des Tastkopfes angepasst. Da die Kanäle A und B identisch sind, wird nur Kanal A beschrieben. Das angeschlossene Eingangssignal kommt über ATTENUATOR A an die VERTICAL SIGNAL CONTROL. In ATTENUATOR A wird der Ablenkkoeffizient bestimmt. Er wird mit Taste AMPL/DIV über den Steuerabschnitt gesteuert (angegeben mit C →).

Mit Taste COUPLING kann ein Softkey-Menü eingeschaltet werden, in dem folgende Funktionen wählbar sind:

- Wahl der Eingangskopplung : AC-DC-GROUND
- Wahl der Eingangsimpedanz: 1 MOhm - 50 Ohm
- OFFSET-Einstellung (mit vorgewählten Werten).

Eine stufenlose Einstellung der Abschwächung ist mit Knopf VARIABLE an der Frontseite möglich; LED UNCAL leuchtet dann.


Die Vertikalverschiebung des Signals wird in VERTICAL SIGNAL CONTROL mit Knopf SHIFT an der Frontplatte eingestellt.

Taste MODE ermöglicht über die Softkeys die Wahl folgender Funktionen:

- INVERT A (B) : Umkehrung des Eingangssignals
- VERTICAL DISPLAY MODES :
 - A : nur Kanal A
 - A und B : Kanäle A und B
 - ADD : A und B addiert
 - B : nur Kanal B
- BW LIMIT : Bandbreitebegrenzung
- MIN/MAX : MIN/MAX-Funktion

Der VERTICAL DISPLAY MODE wird im VERTICAL CHANNEL SWITCH gewählt, der ausserdem Spitzendetektoren enthält, um die mit dem Softkey wählbare MIN/MAX-Funktion zu ermöglichen.

Das gewählte Y-Eingangssignal kommt dann an den CONVERSION AND PROCESSING-Teil.



Trigger section

This oscilloscope can be triggered via signals which are internally derived from channel A or channel B and via two externally applied input signals.

The EXTERNAL trigger signal must be applied to the EXT TRIG. input and is then fed to the TRIGGER SOURCE SELECTOR together with the LINE signal, which is a mains-voltage related trigger signal from the POWER SUPPLY.

In the TRIGGER SOURCE SELECTOR the different trigger sources are selected by the pushbutton (TRIGGER) SOURCE.

When pushbutton SOURCE is pressed, the following trigger sources can be selected via the softkeys:

- A : trigger derived from channel A
- B : trigger derived from channel B
- EXT : trigger derived from EXT TRIG input
- EXT;10 : trigger derived from EXT TRIG input divided by 10
- LINE : trigger derived from the mains voltage

The selected trigger signal is fed to the TRIGGER AMPLIFIER where the trigger coupling, trigger level and trigger slope is determined by respectively the frontpanel controls: COUPLING, LEVEL AND NEGATIVE SLOPE.

The pushbutton COUPLING makes via the softkeys the selection of the following functions possible:

- AC-DC : AC : DC components are blocked.
DC : full trigger bandwidth
- AUTO : The level range is determined by the signal amplitude (peak-peak triggering)
- TVF : Television Frame signal synchronisation
- LF reject : LF trigger signals are rejected (up to 50 kHz)
- HF reject : HF trigger signals are rejected (higher than 50 kHz)
- SLOPE + - : trigger slope selection: positive or negative (can also be selected with button NEGATIVE SLOPE)
- DUAL : triggering on both positive and negative slope of trigger signal

The LEVEL control determines the level of the trigger signal on which an output signal is generated to the ACQUISITION AND CONTROL LOGIC, to start the signal conversion.

The AUTO PEAK-PEAK DETECTOR and the TRIGGER LOGIC are controlled for AUTO LEVEL mode (peak-peak triggering) via the MICROPROCESSOR MANAGEMENT INTERFACE.

For AUTO OFFSET mode the mean value of the trigger signal is fed via the MICROPROCESSOR MANAGEMENT INTERFACE to the microprocessor to calculate the AUTO-OFFSET level.

Trigger moments can be delayed by events which can be applied to the EVENTS/EXT CLOCK input socket.

The LEVEL on which these external EVENT signals pass the EVENTS AND EXTERNAL CLOCK circuit is determined by the LEVEL control via the EVENT LEVEL signal coming from the MANAGEMENT INTERFACE.

These EVENTS are fed to an EVENT COUNTER in the ACQUISITION AND CONTROL LOGIC.

The oscilloscope can operate on an external clock signal, which is applied to the EVENTS/EXT CLOCK input socket. The LEVEL on which the external clock signal passes is determined in the same way as the event level. The external clock signal is fed to the ACQUISITION AND CONTROL LOGIC.

Conversion and processing section

In the conversion and processing section the vertical input signals are digitised, processed and stored in a digital memory. This is done by an Analog to Digital Converter, which receives the vertical input signals via a MULTIPLEXER from the VERTICAL CHANNEL SWITCH AND PEAK DETECTORS.

This MULTIPLEXER is provided with track and hold circuits, which can hold samples that are taken from both channels at the same moment.

The samples are converted then in sequence by the ADC.

Data from the ADC is processed by the DATA PROCESSING CIRCUIT, which performs e.g. the softkey selectable AVERAGE function.

Data from the DATA PROCESSING CIRCUIT is transported to the STORAGE SECTION.

The PEAK DETECTORS, MULTIPLEXER and ADC are controlled by the TIME BASE LOGIC, which on its turn is controlled by the ACQUISITION AND CONTROL LOGIC.

The ACQUISITION AND CONTROL LOGIC performs the following functions:

- Time base setting, controlled by the TIME/DIV switch
- Horizontal mode selection by means of the softkey functions under the HORIZONTAL MODE pushbutton.
 - RECURRENT
 - MULTIPLY
 - SINGLE
 - ROLL
- Trigger delay selection by means of the softkey functions under the TRIGGER DELAY pushbutton.
 - TIME
 - DIV
 - EVENTS
- Selection of EVENTS or EXT CLOCK via the frontpanel EXT CLOCK pushbutton.
- Control of pilot lamps ALIASED, REP ONLY and NOT TRIG'D.

The ACQUISITION AND CONTROL LOGIC is controlled by the microprocessor system, which also controls the DATA PROCESSING circuit.

The DELTA-t circuit is used for the random sampling mode (time base range 100 ns/div - 5 ns/div).

The delta-t output signal is fed to the microprocessor system via the MICROPROCESSOR MANAGEMENT INTERFACE for dot position calculation.

4.4.3 The storage section

After each conversion of a sample into a 10-bits digital code, the code will be stored in a digital memory in the DATA PROCESSING CIRCUIT. The capacity of this memory is 4096 digital values, which is a complete picture of ten horizontal divisions. This memory is configured as a shift register.

When a number of samples is converted and stored in this digital memory, this memory contents is copied into register R0 of the TRACE MEMORY. The number of samples depends on the time base setting. The addresses for the TRACE MEMORY are generated by the ADDRESS GENERATOR under microprocessor system control. It is possible to save the contents of TRACE MEMORY register R0 in one of the other registers R1, R2 or R3 by means of the softkey functions under the SAVE/PLOT pushbutton.

Each of the four registers is able to store 4096 digital 10-bit codes in single channel mode. With both input channels ON, each register capacity is equally divided into 2048 digital 10-bit codes for each input channel.

The TRACE MEMORY can be cleared by pushbutton CLEAR, locked by pushbutton LOCK and enabled for new signal acquisition by pushbutton WRITE.

A TEXT MEMORY is also part of the storage section. In this TEXT MEMORY all display texts are stored under microprocessor control.

Cursor control can be operated via the frontpanel controls 1st and 2nd and the softkey menu under the frontpanel CURSOR pushbutton. It is a microprocessor controlled function which uses the contents of the TRACE MEMORY as input for calculations and the determination of the position of the cursors, and which uses the TEXT MEMORY to store the cursors and the calculation results.

4.4.4 The display section

This section controls the display of the contents of registers R0, R1, R2 and R3 as well as the display of text on the CRT screen under the control of softkey functions. The trace and text data is separated in a vertical and a horizontal component and applied to two signal paths for vertical and horizontal deflection.

Horizontal deflection

Each address of a register corresponds to a specified vertical line of the CRT display along the X-axis; i.e. the display of 10 divisions into 4000 lines.

The address sequence generated by the ADDRESS GENERATOR can be expanded in the HORIZONTAL A versus B / EXPAND circuit and influenced by an additional horizontal position information from the X-POSITION control before they are applied to a HORIZONTAL DIGITAL TO ANALOG CONVERTER.

To provide the discrete steps for the horizontal time base display, the output of the DAC is a linear staircase voltage, which is applied to the horizontal final amplifier via a DOT JOIN / SMOOTH filter and from there to the horizontal deflection plates of the C.R.T.

Via the X-EXPAND DAC an horizontal expand information is applied to the HORIZONTAL FINAL AMPLIFIER for an additional horizontal expand. This information comes from the front panel X-EXPAND control.

Vertical deflection

The contents of each trace register is 4096 10-bit digital codes, each capable of indicating one out of 1024 different signal amplitudes.

These 10-bit digital codes can be inverted or expanded in the VERTICAL INVERT/EXPAND circuit and influenced by an additional position information from the Y-POSITION control before they are converted into the analog representation of the measured input signals by means of a VERTICAL DIGITAL TO ANALOG CONVERTER. From here the signal is applied to the VERTICAL FINAL AMPLIFIER and the vertical deflection plates of the C.R.T. via a VERTICAL DOT JOIN/SMOOTH filter.




Figure 4.56 Display of trace memory contents.

Display functions

With the pushbutton DISPLAY, a softkey menu is activated which makes selection of the following functions possible:

- Register selection
- POSITION selection: assignment of the frontpanel controls Y-POSITION, X-POSITION and X-EXPAND to a selected register and channel.
- Register INVERT
- A versus \square display
- Reduced register text display
- Full register text display

With pushbutton MAGNIFY, a softkey menu is activated for the following functions:

- HORIZONTAL EXPAND $\ast 1 \dots \ast 64$
- VERTICAL EXPAND $Y \ast 5$, $Y \ast 1$ or $Y/5$

Z-control

The Z trace blanking/unblanking signal for the C.R.T. depends on a number of factors, like register selection, A versus B or X-Y display, dot join, text and so on.

The Z-AMPLIFIER is therefore controlled by the microprocessor, a control signal from the vertical and horizontal signal path, the frontpanel INTENSITY controls and the frontpanel FOCUS control.

Plot

Signals PLOT X, PLOT Y and PLOT Z are applied to the ANALOG PLOT OUTPUT circuit and from there to the PLOT OUTPUT socket on the rear panel.

Via the SAVE/PLOT menu a "register plot" function can be selected and via the DISPLAY menu a "screen plot" function can be selected.

Cal

A CAL GENERATOR generates a 1 Vpp - 2 kHz signal for calibration purposes.

4.4.5

The control section

The control signals which are routed to the various circuits are indicated with "C →" for the MANAGEMENT INTERFACE and with "CuP →" for the MICROPROCESSOR.

A MICROPROCESSOR system including a ROM MEMORY for the system program and a RAM MEMORY for the variable data is controlling the instrument.

The following functions are under its control:

- Watching the frontpanel rotary switches
- Reading the actual front panel keys and softkeys and displaying the actual softkey functions on the C.R.T. screen
- Setting of the acquisition circuits via the MICROPROCESSOR MANAGEMENT INTERFACE
- Performance of calculations
- Control of frontpanel pilot lamp
- CRT display control for traces as well as for text
- Plot output control.
- Performance of the AUTO-SET function (with pushbutton AUTO)
- Programming of the front settings (with pushbutton FRONT No)

Microprocessor management interface

All the frontpanel controls, except the CRT CONTROLS, ILLUM, INTENS, INTENS TEXT, TRACE ROT and the POWER ON switch, are activating the various circuits via the microprocessor control circuits and the MICROPROCESSOR MANAGEMENT INTERFACE.

In this MANAGEMENT INTERFACE also various control signals for the signal acquisition are generated under command of the microprocessor.

Options

The pilot lamp REMOTE as well as the pushbutton URQ can be used when an option is installed.

4.4.6

The front section

All the frontpanel rotary switches, pushbuttons, numerical keyboard and softkeys inform the MICROPROCESSOR system about the user's settings via the FRONT CIRCUIT. The MICROPROCESSOR controls the frontpanel pilot lamps via this FRONT CIRCUIT.

4.4.7 The Power supply section

The power supply, which accepts most mains voltage ranges in use (90–264 Vac), produces the various voltages which are used for the electronic circuits, the EHT voltage for the CRT inclusive. It also generates a mains voltage related LINE TRIGGER signal, which is fed to the EXTERNAL TRIGGER INPUT circuit for line triggering.

4.5 BRIEF CHECKING PROCEDURE

4.5.1 General information

This procedure is intended to check the oscilloscope performance with a minimum of test steps and actions required. It is assumed that the operator doing this test is familiar with oscilloscopes and their characteristics.

WARNING: Before switching-on, ensure that the oscilloscope has been installed in accordance with the instructions mentioned in Chapter 3.

NOTE: The procedure does not check every facet of the instrument's calibration; rather, it is concerned primarily with those parts of the instrument that are essential to measurement accuracy and correct operation.
Removing the instrument covers is not necessary to perform this procedure. All checks are made from the outside of the instrument.

If this test is started a few minutes after switching-on, bear in mind that test steps may be out of specification, due to insufficient warm-up time. The check is set up in a logical sequence. The complete flow should be followed carefully to prevent repeating all control settings and input signals at the start of every single check. For a complete check of every facet of the instrument's calibration, refer to the section "Performance Check" in the service manual (for qualified persons only).

No additional test equipment is necessary apart from a 10:1 attenuator probe supplied with the oscilloscope and an adaptor BNC female-probe tip (order nr. 5322 263 50022).

4.5.2 Preliminary settings

- Switch the oscilloscope on and check if the power-up routine is executed.
- Connect the CAL signal from the CAL output socket to the A input via the 10:1 att. probe with automatic range adaption BNC plug.
- Press the green pushbutton AUTO; a square wave should become visible on the CRT screen.
- Set the two INTENS controls for the right intensities for trace and text.
- Adjust the FOCUS screwdriver control for a well defined sharp trace and text.
- Set the ILLUM control for the right intensity of the grid illumination.

4.5.3 Trace rotation

- Press pushbutton VERTICAL COUPLING.
- Press softkey GROUND channel A; a straight line should become visible on the screen.
- Set the trace of channel A in the vertical mid-position of the screen by turning the Y-POSITION control.
- Check that the trace lies in parallel with the horizontal graticule lines; if necessary, readjust the TRACE ROT screwdriver control.

4.5.4 Use of probes

The 10:1 passive probes must be properly compensated before use to avoid pulse distortion or amplitude errors at high frequencies.

For correct adjustment, refer to section 8.1.1.4.

4.5.5 Vertical

As channels A and B are identical, only the procedure for channel A is described.

4.5.5.1 Vertical mode

- Press the green pushbutton AUTO; a square wave with a vertical amplitude of a few divisions becomes visible on the screen.
- Press pushbutton VERTICAL MODE. Check that channel A is selected.
- Press softkey A and B. Channel B is visible as a straight line.
- Press pushbutton VERTICAL COUPLING.
- Press softkey GROUND of COUPLING B.
- Press pushbutton VERTICAL MODE.
- Press softkey ADD; only the square wave is visible.
- Press softkey A. Check the square wave for minimum jump in vertical direction ($< 0,3$ div.).
- Press softkey A-INVERT; the square wave is inverted.
- Press softkey A-INVERT; the square wave is inverted again (normal).
- Press softkey PROCESSING; the VERTICAL PROCESSING menu is visible.
- Press softkey MIN / MAX. Check that the straight horizontal parts of the square wave become more frayed.
- Press softkey MIN / MAX again; MIN / MAX is off.
- Press softkey RETURN; the VERTICAL MODE menu appears.

4.5.5.2 Vertical coupling

- Press pushbutton VERTICAL COUPLING. Check that AC coupling is selected.
- Press softkey AC DC of VERTICAL COUPLING A; DC is selected.
- Check that the displayed square wave is shifted upwards with zero line on vertical midposition.
- Press softkey GROUND of VERTICAL COUPLING A. Check that the input signal is switched off, a straight line is visible on the screen.
- Press softkey AC DC of channel A; DC is selected and the square wave is visible again.
- Press softkey OFFSET of channel A.
- Press softkey UP and DOWN and check if the signal moves up and down.

4.5.5.3 Vertical controls

- Press UP/DOWN control AMPL/DIV on it's left side. Check that the amplitude of the square wave on the screen decreases.
- Press UP/DOWN control AMPL/DIV on it's right side. Check that the amplitude increases.
- Turn VARIABLE left. Check that the amplitude decreases and pilot lamp UNCAL lights up.
- Turn VARIABLE right. Check that the amplitude increases and pilot lamp UNCAL extinguishes.
- Turn SHIFT to the right to move the signal up and to the left to move it down.

4.5.6 Horizontal

- Press the green pushbutton AUTO; a square wave of a few periods becomes visible on the screen.
- Press pushbutton HORIZONTAL MODE. Check that RECURRENT is selected.
- Press softkey SING ARMD; after one sweep the screen should stay steady.
- Press softkey SING ARME again. Check that a new sweep is done.
- Press softkey RECURRENT.
- Press UP/DOWN control TIME/DIV on it's left side. Check that more signal periods appear on the screen.
- Press UP/DOWN control TIME/DIV on it's left side until 200 ms/DIV. Check that pilot lamp ALIASED lights up.
- Press UP/DOWN control TIME/DIV on it's right side until 100 ns/DIV. Check that pilot lamp REP ONLY lights up.

4.5.7 Triggering

4.5.7.1 Trigger source

- Press the green pushbutton AUTO.
- Press pushbutton SOURCE. Check that channel A is selected.
- Press softkey B. Check that the oscilloscope is not triggered; the pilot lamp NOT TRIG'D lights up.
- Press softkey LINE; the pilot lamp NOT TRIG'D extinguishes; the signal is still not stable triggered.
- Press softkey A, the oscilloscope is stable triggered.

4.5.7.2 Trigger coupling

- Press the pushbutton COUPLING. Check that AUTO is selected.
- Press softkey AC DC; AC is selected.
- Turn LEVEL and check if the trigger level indicator at the left of the screen moves up and down.
- Check if the oscilloscope is triggered if the trigger level indicator is between the peaks of the signal.
- Press pushbutton NEGATIVE SLOPE. Check that the trigger slope changes and that the pilot lamp in the pushbutton lights up.
- Press pushbutton NEGATIVE SLOPE again; the pilot lamp extinguishes (+ slope).
- Press softkey AUTO.

4.5.7.3 Trigger delay

- Press pushbutton DELAY.
- Press softkey UP. Check that the signal shifts 1 division to the left.
- Press softkey DOWN twice. Check that the signal shifts two divisions to the right.

4.5.8 Display

- Press the green pushbutton AUTO.
- Turn the X-POSITION control. Check that the signal shifts horizontally.
- Turn the Y-POSITION control. Check that the signal shifts vertically.
- Turn the X-EXPAND control. Check that the signal expands and shrinks and check if the pilot lamp UNCAL functions properly (off in CAL position).

4.5.8.1 Magnify

- Press pushbutton MAGNIFY. Check that *1 is selected.
- Press softkeys Y*5, Y*1 and Y/5 and check if vertical magnify functions correctly.
- Press softkey Y*1.
- Press softkey EXPAND several times until *64 is reached.
- Press pushbutton DOTS. Check that the pilot lamp in the pushbutton lights up and the dots are not joined anymore.
- Press pushbutton SMOOTH and check that the trace becomes a line and the pilot lamp in the pushbutton lights up.
- Press pushbutton DOTS twice; check that the pilot lamps DOTS and SMOOTH are off.
- Press softkey *1.

4.5.9 Memory

- Press pushbutton SAVE/PLOT. Check that the menu appears.
- Press pushbutton CLEAR and check that the screen is cleared and refreshed again.
- Press pushbutton LOCK and check that the screen becomes steady and the pilot lamp in the pushbutton lights up.
- Press pushbutton WRITE and check that the pilot lamp in the pushbutton lights up.

4.5.10 Cursor control

- Press the green pushbutton AUTO.
- Press pushbutton CURSOR and check that the menu appears.
- Press softkey R0 A. Check that two cursors appear on the screen.
- Turn control "1st" and check that the left cursor moves along the signal and that the cursor read out on the C.R.T. is updated.
- Turn control "2nd" and check that the right cursor moves and the read out is updated.

4.5.11 Miscellaneous

- Press pushbutton TRIGGER DELAY.
- Press softkey ENTER.
- Press keys +/- 0 1 2 3 4 and check that the figures appear in the display.
- Press softkey CLEAR.
- Press keys 5 6 7 8 9 and check that the figures appear in the display.
- Press the decimal point key and check that a warning appears in the bottom text area.
- Press pushbutton EXT CLOCK and check that the pilot lamp in the pushbutton lights up and the screen becomes steady.
- Press pushbutton "SETTING FRONT No" and check that the menu appears.
- Press the green pushbutton AUTO to bring the oscilloscope in it's initial state.



5.0 PREVENTIVE MAINTENANCE

5.1 GENERAL INFORMATION

This instrument normally requires no maintenance, since none of its components is subject to wear.

However, to ensure reliable and trouble-free operation, the instrument should not be exposed to moisture, heat, corrosive elements or excessive dust.

5.2 REMOVING THE BEZEL AND THE CONTRAST FILTER

To clean or replace the contrast filter, proceed as follows:

- Push the bezel gently to the right and pull it from the instrument as shown in figure 5.1.
- Remove the contrast filter.
- To prevent scratches, when cleaning the filter, ensure that a clean soft cloth, free from dust and abrasive particles, is used.




Figure 5.1 Removing the bezel and contrast filter.

5.3 REPLACING THE MEMORY BACK-UP BATTERIES

When message

Back up Battery power too low: consult manual

is displayed, the batteries have to be replaced.

To save the settings and traces which are stored in the memory, it is recommended to switch-on the oscilloscope during the replacement of the batteries.

The two 1.5 V penlight batteries (e.g. Philips LR 6 – see also section 6.15) must be installed as described under section 3.3.

NOTE: *It is advisable to remove the batteries when the oscilloscope is stored for longer periods than 24 hours at ambient temperatures below -30°C or above 60°.*

IMPORTANT: **Under no circumstances should the batteries be left in the oscilloscope at ambient temperatures outside the rated range of the battery specifications!**

5.4 RECALIBRATION

From experience, it is expected that the oscilloscope operates within its specifications for a period of at least 1200 hours, or for one year if used infrequently. Recalibration must be carried out by qualified personnel only.

6.0 CHARACTERISTICS

A. General

This instrument has been designed and tested in accordance with IEC publication 348 for Class I instruments.

This specification is valid after the instrument has warmed up for 30 minutes.

Properties expressed in numerical values with tolerances stated, are guaranteed by the manufacturer.

Numerical values without tolerances are typical and represent the characteristics of an average instrument.

Within 5 minutes after switching on, the temperature difference inside the instrument has reached 70 percent of its end value.

B. Contents

- 6.1 Cathode ray tube
- 6.2 Signal acquisition
- 6.3 Channels A and B
- 6.4 Time base
- 6.5 Trigger
- 6.6 Memory
- 6.7 Display
- 6.8 Setting memory
- 6.9 Calculation facilities
- 6.10 Auto setting
- 6.11 Cursors
- 6.12 Calibrator
- 6.13 Power supply
- 6.14 Specification of options
- 6.15 Sundries
- 6.16 Mechanics
- 6.17 Environmental
- 6.18 Safety
- 6.19 Accessories
- 6.20 Optional versions

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.1	CATHODE RAY TUBE		
6.1.1	Type	Philips D18-190GH/129	180 mm rectangular single beam tube.
6.1.2	Useful screen area (h. x w.)	100 mm x 120 mm	For graticule see 6.1.7.
6.1.3	Screen type	GH (P31)	
6.1.4	Total acceleration voltage	16 kV	
6.1.5	Spot size	0,3 mm	Tube only.
6.1.6	Maximum trace distortion		Deviation from straight line.
	-@ screen periphery	1 mm	Outside central 80 mm (vert.) x 100 mm (hor.).
6.1.7	Graticule	Internal, fixed	
	-Illumination	Continuously variable	
	-Size (h. x w.)	80 mm x 100 mm	Centered @ 50 mm from top of CRT screen (hor.) and @ 50 mm from left edge of CRT screen (vert.).
	-Engravings		
	division lines	@ 10 mm	Horizontal and vertical.
	2 mm tick marks	@ 2 mm	On vertical and horizontal central axes.
	0.5 mm tick marks	@ 2 mm	On horizontal lines #2,3,4,6,7,8.
	dots	@ 2 mm	On dotted lines @ 1,5 div and 6,5 div from top of graticule.
	percentages	100-90-10-0 %	To facilitate rise and fall time measurements.
6.1.8	Orthogonality		Measured @ centre of screen.
		90 ± 0,5°	(Angle between X and Y axes, when traces are written in X- and Y direction alternately).
6.1.9	Intensity	Blank to max. intens.	Separate front panel controls for trace and text.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.1.10	Focus	Manually set	Common screwdriver control on front for trace and text.
6.1.11	Trace rotation		Screw driver control on front; direction of screw driver rotation same as direction of trace rotation.
	-Minimum range	14°	Either X or Y trace can be aligned with graticule, when environmental magnetic field is within 0,1 mT.

6.2 SIGNAL ACQUISITION

6.2.1	Sampling type		
	-@ 200 ns/div... 360 s/div:	Real time	
	-@ 5 ns/div... 100 ns/div:	Equivalent time	Random sampling.
6.2.2	Maximum sampling rate		Sampling rate depends on time/div setting.
	-Real time:	250 megasamples/s	
	-Equivalent time:	10 gigasamples/s	Repetitive only.
	-Ext. clock	50 kilosamples/s	Max. aperture uncertainty of 10 us.
6.2.3	Vertical (=voltage) resolution	10 bits	(= 0,1 % of full range).
6.2.4	Horizontal (=time) resolution		
	-In single Ch. or added Ch. acquisition		
	@ 1 ms/div... 5 s/div	4096 samp./acquisition	1 Sample = 0,025 % of full record.
	@ 5 ns/div... 500 us/div	512 samp./acquisition	1 Sample = 0,2 % of full record.
	-In dual channel acquisition		
	@ 1 ms/div... 5 s/div	2048 samp./acquisition	1 Sample = 0,05 % of full record.
	@ 5 ns/div... 500 us/div	512 samp./acquisition	1 Sample = 0,2 % of full record.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.2.5	Record length	10,2 x time/div	Display in unmagnified position.
6.2.6	Acquisition time		
	-Real time	10,2 x time/div	
	5 s/div ...	+ 1 ms/div	}
	1 ms/div		}
			}Exclusive delay time.
	500 us/div ...	+ 10 ms	}
	200 ns/div		}
	-Equivalent time		
	@ 5ns/div	2 s	} After this time there is
	@ 100 ns/div	10 ms	} a 99 % probability of
			} all dots being updated
			} to the new acquisition.
6.2.7	Sources	Channel A	} Both channels can be
		Channel B	} inverted before acquisition.
6.2.8	Acquisition modes	1 channel only	Full memory available for 1 channel.
		2 channels	Simultaneously sampled; 2 channels share memory.
		Ch.A and ch.B added	Full memory available for added channels.
		Average	Combined with 1 channel only, 2 channels or ch. A and ch. B added.
		MIN / MAX	Combined with 1 channel only, 2 channels or ch. A and ch. B added.
6.2.9	Maximum time difference	200 ps	Two 2 channels are sampled simultaneously.

6.3 CHANNELS A AND B

6.3.1	Input connector	BNC with probe read-out	Probe read-out causes instrument to change V/div indication, input impedance and attenuator setting according to probe (when fitted with a probe indicator).
-------	-----------------	-------------------------	--

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.3.2	Input impedance (In high Z position)		For frequency > 1 MHz see Fig. 6.1.
	-R parallel	1 M Ohm \pm 1 %	}In DC position of input }coupling. In AC position }of input coupling: 18 nF
	-C parallel	14 pF	}In series with R par. & }C par. }In 0 position of input }coupling: R par. = ∞
	-Maximum input capacitance difference	1,5 pF	Capacitance difference between channel A, channel B and/or trigger input.
6.3.3	Input impedance (In 50 ohm position)		
	-R parallel	50 Ohm \pm 1 %	In DC, AC and 0 position of input coupling.
	-VSWR (typical)	1,2:1	@ 200 MHz; in AC and DC pos. of input coupling.
6.3.4	Input coupling	a.c. d.c. 0	In 0 position: channel input disconnected from BNC and connected to ground.
6.3.5	Max. input voltage		Instrument should be properly grounded through the protective-ground conductor of the power cord.
	-In high Z position (d.c. + a.c. peak)	300 V	Up to 1 MHz; for freq. > 1 MHz see Fig. 6.2.
	-In 50 ohm position d.c.	5 V	}
	a.c. (r.m.s.)	5 V	} Max. 50 mJ during any
	a.c. (peak)	50 V	} 100 ms interval. }
6.3.6	Deflection coefficient		
	-Steps	5 mV/div...5 V/div	In a 1-2-5 sequence of 10 steps.
	-Vernier ratio	1:2,5	Continuously variable between steps.
	Read out accuracy	\pm 15 %	
	-Error limit (Ambient 15...35°C)		Add 3 % for ambient 0...50°C.
	overall	\pm 2 %	}
	up to memory	\pm 1 %	}Vernier in 0 position. }
	additional error	\pm 15 %	Vernier not in 0 position.




	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.3.7	Dynamic range	10 div	
6.3.8	D.c. offset control		Related to BNC input.
	-Range ($\pm 5\%$)		Input voltage within the limits of 6.3.5.
	Att. @ 5 mV/div ...20 mV/div	$\pm 5 \text{ V}$	}In AUTO OFF-SET, the }offset is automatically }controlled such, that }average d.c. level of }signal is presented at }screen centre (\pm }2 div), provided signal }is within offset range. }Shift is set to zero } (mid) level.
	Att. @ 50 mV/div ...0,2 V/div	$\pm 50 \text{ V}$	
	Att. @ 0,5 V/div ...5 V/div	$\pm 300 \text{ V}$	
	-Resolution ($\pm 5\%$)		
	Att. @ 5 mV/div ...20 mV/div	5 mV	
	Att. @ 50 mV/div ...0,2 V/div	50 mV	
	Att. @ 0,5 V/div ...5 V/div	0,5 V	
6.3.9	Shift range	$\pm 5 \text{ div}$	From screen centre.
6.3.10	Frequency response (in 50 ohm position)		Z source: 50 Ohm
	-Lower transition point of BW		
	Input coupling in DC position	d.c.	
	Input coupling in AC position	$\leq 10 \text{ Hz}$	
	-Upper transition point of BW (Ambient: 15..35°C)	$\geq 200 \text{ MHz } (-3 \text{ dB})$	Deviation max. 30 MHz for ambient: 0...50°C.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.3.11	Freq. resp. (In hi.Z pos. through probe) -Lower transition point of BW Input coupling in DC position Input coupling in AC position -Upper transition point of BW (Ambient: 15...35°C)	 d.c. $\leq 1 \text{ Hz}$ $\geq 200 \text{ MHz } (-3 \text{ dB})$	 Z source = 25 Ohm. Probe according to 6.19. Deviation max. 30 MHz for ambient: 0...50°C.
6.3.12	Bandwidth limiter -Starting point of HF rejection -Slope	 20 MHz (-3 dB) 6 dB/octave	
6.3.13	Pulse response (In 50 Ohm position) (exclusive first dot after transient) -Rise time (Ambient: 15...35°C) -Pulse aberrations Overshoot Ringing	 $\leq 1,75 \text{ ns}$ } $\leq 6\%$ }	 Z source = 50 Ohm; measured over central 6 div. (Calculated from bandwidth x Rise time = 0,35). Add max. 0,25 ns for ambient: 0...50°C. Tested with ca 1 ns rise time pulse. During first 10 ns after transient.
6.3.14	Pulse resp. (in hi.Z pos. through probe) (exclusive first dot after transient) -Rise time (Ambient: 15...35°C) -Pulse aberrations Overshoot Ringing	 $\leq 1,75 \text{ ns}$ } $\leq 6 \%$ }	 Z source: = 50 Ohm; measured over central 6 div. Probe according to 6.19. (Calculated from bandwidth x Rise Time = 0,35). Add max. 0,25 ns for ambient: 0...50°C. Tested with ca 1 ns rise time pulse. During first 10 ns after transient.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.3.15	Max. base line instability		
	-Jump (Ambient 15...35°C):		Add 25 % for ambient 0...50°C.
	when switching to added mode	0,3 div	
	when switching to MIN / MAX mode	0,5 div	
	between any V/div positions	0,15 div	
	when actuating inverter switch	0,3 div	
	between AC, 0 and DC position	0,1 div	
	when rotating vernier	0,6 div	Input externally grounded.
	between any time/div positions	0,6 div	
	when switching to magn. x5	0,5 div	
	-Drift	0,1 div/h	} Measured in 20 mV/div position. }
	-Temperature coefficient	$\pm 0,05$ div/K	
6.3.16	Common mode rejection ratio		Both channels @ same attenuator setting; vernier for V/div setting adjusted for optimal CMMR at 10 kHz measured with max. 1 div input signal on each channel, (± 4 div around zero).
	-@ 1 MHz	100:1	
	-@ 50 MHz	20:1	
6.3.17	MIN / MAX function		Time base setting 5 μ s/div ... 360 s/div. Average switched off.
	-Accuracy > 50 %	@ pulse > 3 ns	
	-Reset time	20 ns	

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.3.18	Average		Average formula after the first front change (MIN / MAX switched off). $So(n) = So(n-1) + \frac{Sn - So(n-1)}{C}$
	Constant is max.	64x 32 x or off	In ROLL mode.
6.3.19	Cross talk (according to IEC 351)	<-30 dB @ 100 MHz <-50 dB @ 2 MHz	
6.4	TIME BASE		
6.4.1	Modes	Recurrent Single shot Single scan Multiple shot Multiple scan Roll	Up to 4 shots. Up to 4 scans. Can be stopped manually or by trigger.
6.4.2	Time coefficients		
	-In recurrent	5 ns/div...5 s/div	
	-In single scan and multiple scan	5 ns/div...100 ns/div	
	-In single shot and multiple shot	200 ns/div...5 s/div	
	-In roll mode	50 ms/div...360 s/div	
	-With EXT CLOCK	Depending on clock frequency	Input via EXT CLOCK, every clock pulse = sample is taken, so for single channel 4k samples are stored and for dual channel 2 times 2k samples.
	-Error limit (Ambient 15...35 °C)		
	In equivalent time mode	± 4 %	Add 2 % for ambient: 0...50°C.
	In real time mode	± 1 %	Add 0,5 % for ambient: 0...50°C.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.5	TRIGGER		
6.5.1	Sources		
	-Signal trigger	Channel A Channel B EXT LINE	
	-Events	EXT	Serves as delay to signal trigger.
6.5.2	Input connectors	BNC	
6.5.3	Input impedance of EXT trigger inputs		
	-R parallel	1 M Ohm \pm 1 %	In DC position of input coupling.
	-C parallel	14 pF	
	-Max. input capacitance difference	1,5 pF	Difference between channel A, channel B and EXT trigger input.
6.5.4	Coupling		
	-Signal trigger	d.c. a.c. LF rejected HF rejected Auto level TVF	According to CCIR.
	-Clock & events trigger level	TTL ECL Through variable level	Adjustable via menu "trigger coupling" - events.
6.5.5	Max. input voltage (d.c. + a.c. peak)		Instrument should be properly earthed through the protective-earth conductor of the power cord.
	-Clock trigger	300 V	

CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.5.6	Signal trigger sensitivity (Ambient 15..35°C)	
-Channel A or B		
@ 300 MHz	≤ 3 div	
@ 200 MHz	≤ 1 div	
@ 30 MHz	$\leq 0,5$ div	
-EXT		
@ 300 MHz	≤ 300 mV	
@ 200 MHz	$\leq 0,1$ V	
@ 30 MHz	$\leq 0,05$ V	
-EXT/10		
@ 300 MHz	≤ 3 V	
@ 200 MHz	≤ 1 V	
@ 30 MHz	$\leq 0,5$ V	
6.5.7	Slope selection	
	Positive going Negative going Dual slope	Level adjustable; not effective in random sampling and TVF.
6.5.8	Signal level control range	
-Channel A or B	± 8 div	} When not in AUTO position of trigger mode. }
-EXT	$\pm 0,8$ V	
-EXT/10	± 8 V	
-Any source	Related to peak value	} In AUTO position of trigger mode.
6.5.9	Frequency response	Trigger BW not affected by bandwidth limiter.
-Lower transition point of BW		Ch. A and Ch. B coupling cascaded with trig. coupl.
Trigger coupling in DC position	d.c.	
Trigger coupling in AC position	10 Hz (- 3dB)	
Trigger coupling in LF reject pos.	50 kHz (- 3 dB)	
-Higher transition point of BW	see also 6.5.6.	
Trigger coupling in HF reject pos.	50 kHz (-3 dB)	

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.5.10	Trigger delay		
	-Range	-10...9999 div	}can also be indicated in }time. } }
	@ 5 ns/div ... 100 ns/div	-10...500 div	
	-Number of events	1...9999	
	Max. frequency	5 MHz	
	-Accuracy	$\pm 1\%$	
	@ 5 ns/div ... 100 ns/div	$\pm 4\%$	
6.6	MEMORY		
6.6.1	Memory size		
	-Registers	4	Registers #0, #1, #2, #3
	-Register depth	4096 words	
	-Wordlength	10 bits	
6.6.2	Functions	Clear	Register #0 is cleared, incl. pre-trigger memory and blanked if DOTS is selected.
		Save	Contents of register #0 is saved in selected register (#1, #2 or #3).
		Write	Acquired signal is written into register #0.
		Lock	Memory system is locked, including register #0.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.7	DISPLAY		
6.7.1	Sources	Register #0 Register #1 Register #2 Register #3	} }In any combination. } }
6.7.2	Display expansion		
	-Horizontal		
	Steps:	1x...64x 1x...8x	Y versus t A versus B
	Vernier ratio	1:2.2	Continuously variable between steps. Recalculated value is displayed with an accuracy of $\pm 5\%$.
	-Vertical		
	Steps:	0.2x, 1x and 5x	Both in Y versus t and Y versus X modes.
6.7.3	Display manipulations	Smooth	Reduces noise by adding a filter in the display section; that is only effective at time base 500 us/div ... 360 s/div.
		Dot join	Linearly interpolated between measured dots.
		Invert	All registers can be inverted.
6.7.4	Position range		All channels can be positioned independently.
	-Horizontal	± 5 div	From screen centre.
	-Vertical	± 5 div	From screen centre.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.8	SETTING MEMORY		
6.8.1	Memory size	Max. 77 front settings	
6.8.2	Functions		
	Save		Actual settings are stored in memory, replacing contents of memory cell indicated on CRT.
	Insert		Actual settings are stored in memory; insertion is after memory cell indicated on CRT.
	Delete		Contents of memory cell indicated on CRT is deleted.
	Recall		Actual settings are replaced by contents of memory cell indicated on CRT. Actual settings are saved in "back-up" memory (= mem. cell #0).
	(Recall) Next		Actual settings are replaced by contents of memory cell indicated on CRT increased by 1. Actual settings are saved in "back-up" memory (= memory cell #0).
	(Recall) Previous		Actual settings are replaced by contents of memory cell indicated on CRT decreased by 1. Actual settings are saved in "back-up" memory (= memory cell #0).

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.9	CALCULATION FACILITIES		
6.9.1	Functions	RMS value Mean value Peak to peak value Rise or fall time Frequency (1/dt) Multiplication	} } of portion between } cursors, or markers if } LOCATE is choosen. } Whole register.
6.10	AUTO SETTING		
6.10.1	Settling time	Typical 3 s	During plot, AUTO SET is not possible.
6.10.2	CRT functions		
	-Focus	Not influenced	
	-Trace intens	Not influenced	
	-Text intens	Not influenced	
6.10.3	Display functions		
	-Select	To register #0	
	-X-position	Zero	
	-Y-position	Zero	
	-Invert	Off	Only for register #0.
	-X-expand	*1	Vernier calibrated.
	-Y-expand	*1	
	-A versus B	Off	Only for register #0.
6.10.4	Cursors	Off	
	-Calculation	Off	
6.10.5	Text		
	-Reduced	Off	
	-Bottom text lines	Off	

CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.10.6 Vertical acquisition		
Y-deflection source	Every source having a triggerable signal at its input	}Channel A if no trigger is found. }
Input impedance		
-Accessory with probe read out	According to probe read out	
-Otherwise	Not affected by AUTO SET	
Input coupling	ac	
Y-deflection		Each channel is independently set.
-10 mV < input voltage < 30 V	Approx. 4 div	}Vernier in calibrated position. }
-Input voltage < 10 mV	Channel at 200 mV/div	} }
		Due to trigger uncertainty at freq. > 60 MHz or at duty cycle < 50% sensitivity can deviate from above, but signal will remain on the screen.
Channel inverter	Off	
-Add	Off	
-MIN / MAX	Off	
-Bandwidth limiter	Off	
-Average	Off	
-Offset	Zero	
Y base line position		
-In single channel display	Centre of screen $\pm 0,3$ div	
-In dual channel display		
Ch. A	$2 \pm 0,3$ div above centre screen	
Ch. B	$2 \pm 0,3$ div below centre screen	

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.10.7	Horizontal acquisition		
	Mode	Recurrent	
	External clock	Off	
	TB deflection coefficient		For TVF not affected.
	-Signal frequency 40 Hz ... 80 MHz	Min. 2, max 6 signal periods over 10 div.	
	-Signal freq. > 80 MHz	5 ns/div	
6.10.8	-When no trigger found	2 ms/div	
	Triggering		
	-Delay	Off	For TVF not affected.
	-Events	Off	For TVF not affected.
	-Source		
	Triggerable signal @ ext input	Ext	
	No signal @ ext input, but trigg. signal @ channel A or B	Channel A or B	Channel with lowest input frequency is selected. (Channel A when frequencies are equal).
	No triggerable signal @ any input	Channel A	
	-Mode	Auto	For TVF not affected
		TVF	Trigger on fieldpulse with CCIR TV system.
	-Level	50 ... 70% of peak to peak value	Dc component of signal neglected.
	-LF reject	Off	
	-HF reject	Off	
	-Slope	Positive	For TVF not affected.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.11	CURSORS		
6.11.1	Cursor intensity control	Independent of trace intensity but combined with text intensity	
6.11.2	Horizontal resolution		
	-In single channel mode	1 : 4096	
	-In dual channel mode	1 : 2048	
6.11.3	Vertical resolution	1 : 1024	
6.11.4	Read out resolution	3 digits	
6.11.5	Voltage cursors		
	-Error limit < Ambient: 15..35°C	$\pm 2 \%$	Referred to input at BNC, error of probes etc. excluded. Add 3 % for ambient: 0...50°C.
	-Cursor Range	Visible part of signal	Cursors cannot pass each other, (to avoid negative readings).
6.11.6	Time cursors		
	-Error limit	$\pm 0,2 \%$	
6.12	CALIBRATOR		
6.12.1	Wave form		
	-Shape	Square wave	
6.12.2	Internal impedance	50 Ohm $\pm 1 \%$	
6.12.3	Output voltage (peak to peak)	1 V $\pm 1 \%$	Open voltage; (halves when terminated with 50 Ohm). Positive going with respect to ground.
6.12.4	Output current (peak to peak)	20 mA $\pm 2 \%$	Output short circuited; (halves when terminated with 50 Ohm).
6.12.5	Frequency	2 kHz	

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.13	POWER SUPPLY		
6.13.1	Source voltage a.c. (r.m.s)		
	-Nominal	100 V ... 240 V	
	-Limits of operation	90 V ... 264 V	
6.13.2	Source frequency		
	-Nominal	50 Hz ... 400 Hz	
	-Limits of operation	45 Hz ... 440 Hz	
6.13.3	Source waveform characteristics		@ Nominal source voltage.
	-Max. waveform deviation factor	10%	
	-Crest factor	1,27...1,56	
6.13.4	Allowable power source interrup- tion	At least 20 ms	@ Nominal source voltage. After this time oscillo- scope settings are saved before instrument goes down. Automatic power up after restoration of Power line voltage. (For setting retention: see 6.15.1).
6.13.5	Power consump- tion (a.c. source)		
	-Nominal	160 W	

6.14 OPTIONS

For specification of an option refer to the separate manual of the option.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.15	SUNDRIES		
6.15.1	Data and settings retention		When instrument is switched off or during line power failure.
	-Memory back up voltage	2 V ... < 3,5 V	
	-Memory back up current drain	Typical 12 uA	@ 25°C.
	-Recommended Batteries:		
	type	LR 6	According to IEC 285, (= Alkaline manganese penlight battery), e.g. PHILIPS LR6 (9299 000 20734) or DURACELL MN 1500.
	quantity	2 pcs	
	-Temperature rise of batteries	20 K	After warming up period of instrument
	-Retention time	Typical 2 years	@ 25°C, with recommended (fresh) batteries.
	-Temperature Range	0...+70°C	@ -40...0°C settings retention is uncertain. It is advised to remove batteries from instrument when it is stored during longer period (> 24 h) below -30°C or above 60°C. UNDER NO CIRCUMSTANCES BATTERIES SHOULD BE LEFT IN THE INSTRUMENT @ TEMPERATURES BEYOND THE RATED RANGE OF THE BATTERY SPECIFICATIONS!

CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.15.2 Probe Read Out		With Philips probe provided with indicator ring.
-Input impedance setting		
Passive high impedance probe	}Fixed @ 1 MOhm }	
Passive 50 Ohm probe	}	
Passive 50 Ohm attenuator	}	
Passive 75 Ohm in 50 Ohm out	}Fixed @ 50 Ohm }	
attenuator	}	
Active current probe	}	
Active isolation probe	}	
-Vertical sensitivity setting		
Passive probe	Not affected	
Passive attenuator	Not affected	
Active current probe	20 mV/div	Can be manually changed.
Active isolation probe	20 mV/div	Can be manually changed.
-V/div and voltage cursor read out		
Passive 10:1 probe	10 x higher	}
Passive 10 x attenuator	10 x higher	}
Passive 100:1 probe	100 x higher	}Offset read out is }changed accordingly.
Passive 100 x- attenuator	100 x higher	}
Active current probe	In divisions	}
Active isolation probe	In divisions	}

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.15.3	Analog plot output		
	-Connector	DIN 5 pole 45°	
	-Functions	Screen dump	Included channel identifier.
		Memory dump	Register selectable.
	-Sensitivity	1 V/Full screen ± 3 % 1 V/Full memory ± 3 %	} } Horizontal and vertical. } }
	-Pen lift	TTL compatible	Pen-up is selectable (0 or 1). Open collector output; max. 12 V.
	-Plot time per dot	20 ms .. 2 s	Software selectable. Signal dependent.
	-Plot sequence	Channel A first	In dual channel operation; with more registers starting with the lowest number.
6.16	MECHANICS		
6.16.1	Height		Fits 5E in 19 inch rack.
	-Without feet and accessory pouch	176 mm (6,9 in.)	Add 15. mm (0,6 in.) for feet.
	-Feet and accessory pouch included	250 mm (9,8 in.)	
6.16.2	Width	419 mm (16,5 in.)	Add 46 mm (1,8 in.) for handle.
6.16.3	Depth		
	-Handle excluded	570 mm (22,5 in.)	Add 35 mm (1,4 in.) for protective front cover.
	-With extended handle	670 mm (26,4 in.)	
6.16.4	Mass	18 kg	
6.16.5	Operating position	Horizontal	Standing on feet; may be tilted by handle.
6.16.6	Finish	Epoxy powder coated	
6.16.7	Printed circuit boards	Glass laminate epoxy	
6.16.8	Cooling	Fan aided convection	Maintenance free.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.17	ENVIRONMENTAL		
6.17.1	General		
	The characteristics are valid only if instrument is checked in accordance with the official checking procedures. Details on these procedures and failure criteria are supplied on request.		
6.17.2	Meets environmental requirements of	MIL-T-28800C Type III Class 5, Style D	
6.17.3	Temperature.		Memory back-up batteries removed from instrument, unless batteries meet temperature specifications (see also 6.15.1).
	-Operating		
	Min. low temperature	0°C	Cf. MIL-T-28800C par. 3.9.2.3 tested cf. par. 4.5.5.1.1.
	Max. high temperature	+ 50°C	Cf. MIL-T-28800C par. 3.9.2.4 tested cf. par. 4.5.5.1.1.
	-Non operating (Storage)		
	Min. low temperature	-40°C	Cf. MIL-T-28800C par. 3.9.2.3 tested cf. par. 4.5.5.1.1.
	Max. high temperature	+ 75°C	Cf. MIL-T-28800C par. 3.9.2.4 tested cf. par. 4.5.5.1.1.
6.17.4	Maximum humidity		Cf. MIL-T-28800C par. 3.9.2.2 tested cf. par. 4.5.5.1.1.
	-Operating and non-operating (Storage)	95% Relative humidity	
6.17.5	Maximum altitude		Cf. MIL-T-28800C par. 3.9.3 tested cf. par. 4.5.5.2.
	-Operating	4,5 km (15000 feet)	Memory Back-up batteries removed from instrument, unless batteries meet maximum altitude specs.
	-Non-operating (storage)	12 km (40000 feet)	Maximum operating temperature derated 3°C for each km (for each 3000 feet) above sea level.

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.17.6	Vibration (Operating)		Cf. MIL-T-28800C par. 3.9.4.1 tested cf. par. 4.5.5.3.1.
	-Freq. 5...15 Hz		
	Sweep time	7 min	
	Excursion (pk to pk)	1,5 mm	
	Max acceleration	7 m/s ² (0,7 x g)	@ 15 Hz.
	-Freq. 5...25 Hz		
	Sweep time	3 min	
	Excursion (pk to pk)	1,0 mm	
	Max acceleration	13 m/s ² (1,3 x g)	@ 25 Hz.
	-Freq. 25...55 Hz		
	Sweep time	5 min	
	Excursion (pk to pk)	0,5 mm	
	Max acceleration	30 m/s ² (3,0 x g)	@ 55 Hz.
	-Resonance dwell	10 min	@ each resonance freq. (or @ 33 Hz if no resonance was found). Excursion cf. 6.17.6.
6.17.7	Shock (Operating)		Cf. MIL-T-28800C par. 3.9.5.1 tested cf. par. 4.5.5.4.1.
	-Amount of shocks		
	total	18	
	each axis	6	(3 in each direction).
	-Shock wave form	half sine wave	
	-Duration	11 ms	
	-Peak acceleration	300 m/s ² (30 x g)	
6.17.8	Bench handling		Cf. MIL-T-28800C par. 3.9.5.3 tested cf. par. 4.5.5.4.3.
	-Meets require- ments of:	MIL-STD-810 methode 516, proced. V	
6.17.9	Salt atmosphere		Cf. MIL-T-28800C par. 3.9.8.1 tested cf. par. 4.5.6.2.1.
	-Structural parts meet requirements of:	MIL-STD-810 methode 509, proced. I salt solution 20%	

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.17.10	EMI (Electro magnetic interface) meets requirements of:	MIL-STD-461 Class B VDE 0871 and VDE 0875 Grenzwertklasse B	Applicable requirements of Part 7: CE03, CE07, CS01, CS02, CS06, RE02, RS02, RS03.
6.17.11	Magnetic radiated susceptibility -Maximum deflection factor	7 mm/mT (0,7 mm/gauss)	Tested conforming IEC 351-1 par. 5.1.3.1. Measured with instrument in a homogeneous magnetic field (in any direction with respect to instrument) with a flux intensity (peak to peak value) of 1,42 mT (14,2 gauss) and of symmetrical sine wave form with a frequency of 45...86 Hz.
6.17.12	Packing Meets requirements of:	NLN-L88	
6.17.13	Transportation Meets requirements of: -Packaged transportation drop meets requirements of: -Packed transportation vibration meets requirements of:	AN-D628 Mat. safe transp. ass. procedure 1A-B-2 Mat. safe transp. ass. procedure 1A-B-1	
6.18	SAFETY		
6.18.1	Meets requirements of:	IEC 348 Class I VDE 0411 Class I UL 1244 CSA 556B	
6.18.2	Approvals	VDE 0411 (applied for) UL 1244 (applied for) CSA 556 (applied for)	

	CHARACTERISTIC	SPECIFICATION	ADDITIONAL INFORMATION
6.19	ACCESSORIES		
6.19.1	Accessories furnished with instrument	2 x PM8929/09 Blue contrast filter Operating manual Front cover	10 MOhm, 10:1 passive probe with read out (1,5 m). Factory installed.
6.20	OPTIONAL VERSIONS		
6.20.1	General		These options can be factory installed only.
6.20.2	Power Cord	Universal european North american United kingdom Swiss Australian	}Length 2,1m , (82,7 in.). }VDE, KEMA listed. }CSA, UL listed. }BSI listed. }SAV listed. }SAA listed.
6.20.3	Cabinet	Rack mount	
6.20.4	IEEE 488/RS 232C interface	PM8956	




Figure 6.1. Input resistance R_{par} and capacitance C_{par} versus frequency.




Figure 6.2. Maximum input voltage (peak to peak) derating versus frequency.

7.0 PM3320 VERSIONS – ADDITIONAL INFORMATION

The version of your oscilloscope is indicated on the type plate situated on the rear panel (see fig.7.1.).

The version is indicated as follows:

- 1 PM3320/XY :in type number
- 2 12nc: 9444 W33 20XYZ :in code number

WXYZ are represented by numbers.

These numbers are given in this section and each version is briefly described.




Figure 7.1 Rear panel with indicated type plate position.

Typenumber: PM3320/XY

Codenummer: 9444 W33 20XYZ

W=0:

Standard version

XY=00

Basic instrument without options.

XY=30

Basic instrument without options but with 19 inch rackmount parts.

XY=40

Basic instrument with IEEE 488/RS232-C option installed.

XY=80

Basic instrument with IEEE 488/RS232-C option installed and with 19 inch rackmount parts.

W≠0:

Non standard version


Z:

Power-cord-version indication


On type plate on instrument: Z always ■

In the codenumber on the packing in which the instrument was shipped the indication is as follows:


Z=1 Standard European version
220 V / 16 A / 50 Hz




Z=3 U.S.A. version
(U-version)
120 V / 15 A / 60 Hz



Z=4 United Kingdom (UK) version
including line cord plug
fuse of 13A (type C)
240 V / 13 A / 50 Hz



Z=5 Swiss version
220 V / 10 A / 50 Hz



Z=8 Australian version
240 V / 10 A / 50 Hz




Figure 7.2 Mains connectors.

8.0 ACCESSORY INFORMATION

8.1 ACCESSORIES SUPPLIED WITH THE INSTRUMENT

8.1.1 Passive probe PM8929/09 with automatic range indication

8.1.1.1 Introduction

This 10 x attenuator probe is provided with a special BNC plug with built-in resistor for automatic range indication to advance the V/DIV reading by 10x.

The probe consists of 3 separate units:

- a compensation box having a BNC male connector output,
- a cable assembly,
- a probe body including probe tip and RC assembly.

At delivery the probe has been adjusted to an oscilloscope with an input capacitance of 12pF.

8.1.1.2 Characteristics

- Properties expressed in numerical values with tolerances stated, are guaranteed by the manufacturer.
- Numerical values without tolerances are typical and represent the characteristics of an average probe.

Electrical

Note: *These characteristics are valid with a termination of 1 MOhm oscilloscope input, unless otherwise stated.*

Designation	Specification	Additional information
Attenuation (d.c.)	10 x + or - 2 %	
Input impedance		
-parallel resistance		
at d.c.	10 MOhm + or - 1,5 %	
at a.c.	see fig. 8.1.	
-parallel capacitance	13,5 pF	(for parallel capacitance as function of frequency, see figure 8.1).
up to 100 kHz		
Compensation range		
-input capacitance of oscilloscope	5 pF ... 20 pF	

Bandwidth

-probe only bandwidth at osc.
input cap. 10 pF
or less

d.c. ... 450 MHz
(-3 dB)

>10 pF

d.c. ... 300 MHz
(-3 dB)

-max. useful system
bandwidth at osc.
input cap. of 10pF

d.c. ... 300 MHz
(-3 dB)

Note: Up to this freq.
the system (probe + osc.)
bandwidth is $\geq 95\%$ of
the "osc. only
bandwidth".

Pulse response

Aberrations in addition
to osc. aberrations.
Oscilloscope bandwidth
 \leq useful bandwidth.

-Overshoot

$< 6\%$

-Ringing during
first 30 ns after
leading edge

+ or - 5 %, or
7 % pk - pk

-Ringing there-
after

+ or - 2 %

-Tilt

$\leq 2\%$

Signal delay

7,6 ns + or -
200 ps

Measured between tip to
BNC-output connector.

Maximum voltage

-max. non destruc-
tive input voltage
(d.c. + a.c. peak)

500V

0...2 MHz approx. for
derating see figure 8.2.

-test voltage
(d.c.): type test

2,42 kV

During 1 min (resistance
value adapted to test)

-performance check

2,42 kV

During 1 sec

Mechanical

-Dimensions

length

width

height

probe body

57 mm

14 mm(max)

cable assy

1500 mm

■ mm(max)

compensation box

38 mm

16 mm

15 mm BNC excluded

pouch

275 mm

195 mm

-Mass

137 g

Standard probe with
accessories in pouch.

Environmental

The characteristics are valid only if the instrument is checked in accordance with the official checking procedure. Details on these procedures and failure criteria are supplied on request by the PHILIPS-organisation in your country, or by PHILIPS, INDUSTRIAL AND ELECTRO-ACOUSTIC SYSTEMS DIVISION, EINDHOVEN, THE NETHERLANDS.

Operating temperature	-10 °C ... +55 °C
Storage temperature	-62 °C ... +85 °C
Maximum humidity	95 % relative humidity
Altitude	
-operating	To 4500 m
-non-operating	To 12000m
Vibration (operating)	
- freq. 5...15 Hz	7 min each axis, excursion 1,5 mm (p-p) and 7 m/s ² (0,7 g) acceleration at 15 Hz.
- freq. 15...25 Hz	3 min each axis, excursion 1 mm (p-p) and 13 m/s ² (1,3 g) acceleration at 25 Hz.
- freq.25...55 Hz	5 min each axis, excursion 0,5 mm (p-p) and 30 m/s ² (3 g) acceleration at 55 Hz.
Resonance dwell	10 min at each resonance freq.
Shock (operating)	300 m/s ² (30 g), half sine-wave shock, duration is 11 ms.(3 shocks per direction for a total of 18 shocks).




Figure 8.1. Input impedance v.s. Frequency.




Figure 8.2. AC component (pk) of max. input voltage v.s. Frequency.

Accessories

-Accessory kit, contents:

- Earth cable
- Spring-loaded test clip
- Set marking rings
- Probe tip (2x)
- Insulating cap
- DIL cap
- Wrap pin adapter
- Earth bus

- Instruction manual

8.1.1.3 Description of accessories

Earth cable: To minimize ringing in a signal, an earth cable is provided. This cable must first be plugged onto the probe body and then be connected to the nearest earth point of the circuit to be measured.

Spring-loaded test clip: This is a provision for hands-free connection to a test point or component lead.

Marking rings: At delivery a set of 3 different colour marking rings (red, white and blue) are provided. This can be used to help identify the specific probes when using more than one probe on an oscilloscope.

Probe tip: A spare set of 2 probe tips are standard supplied with the probe. When a probe tip is damaged it can be pulled out by means of a pair of pliers. Then a new tip must be firmly pushed in.

Insulating cap: An insulating cap is provided to cover the metal part of the probe during measurements in densely wired circuits.

D.I.L. cap: This is a cap facilitating measurements on dual-in-line integrated circuits.

Wrap pin adapter: The wrap pin adapter is a provision to make hands-free connection to a wire wrapped pin circuit.

Earth bus: This is a provision to minimize ringing in VHF signals, when earthing must be as short as possible.

8.1.1.4 Adjustments

Matching the probe to your oscilloscope

The measuring probe has been adjusted and checked by the manufacturer. However, to match the probe to your oscilloscope, the following manipulation is necessary.

- Connect the measuring pin to the CAL-socket of the oscilloscope.
- A trimmer can be adjusted through a hole in the compensation box to obtain optimum square-wave response, see figure 8.3, 8.4 and 8.5.




Figure 8.3 Over-compensation




Figure 8.4 Correct compensation




Figure 8.5 Under-compensation

8.1.2 Blue contrast filter




Figure 8.6 Blue contrast filter. (Factory installed !)

8.1.3 Front cover




Figure 8.7 Oscilloscope with front cover.

8.2 OPTIONAL ACCESSORIES

8.2.1 IEEE 488/RS232-C bus intelligent interface PM 8956

The IEEE 488/RS232-C is a general-purpose bus interface designed according to the IEEE 488/RS232-C standard. This option can be either retrofitted or factory installed. It enables the oscilloscope to be used in a measuring system together with other IEEE 488/RS232-C bus compatible instruments.

For installation instructions and detailed operating and programming information concerning this facility, refer to the programming manual.